



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Peter Lyskowski
 Director

Jeremiah W. (Jay) Nixon
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP2222 **Manufacturer:** Guth
Model Number: 12V500
Agency: MSHP (GHQ)
Agency Address: 1510 E ELM ST, JEFFERSON CITY, MO 65101

NIST THERMOMETER INFORMATION

Serial Number: 306168 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 8/13/2015 **Date of Expiration:** 8/13/2016

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
--------------------------	---------------------	-----------------------------

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.00	0.025

The combined uncertainty is calculated with a k=2 value.

Date of testing: 2/26/2016
Certification Expiration: 2/26/2017
Simulator testing technician: J. CLEVELAND

Notes on Condition: STATUS 8 ON STARTUP

Deviation(s) from method: none

DHSS BAP Scientist Approving: ELLEN STRAWSINE

Certification No: MP2222_2262016

X

DHSS BAP Scientist Approving