



Missouri Department of Health and Senior Services
P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Peter Lyskowski
Director

Jeremiah W. (Jay) Nixon
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2252 **Manufacturer:** Guth
Model Number: 10-4D
Agency: HOLDEN PD
Agency Address: 110 W 3RD ST., HOLDEN, MO 64040

NIST THERMOMETER INFORMATION

Serial Number: 358440 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 9/8/2015 **Date of Expiration:** 9/8/2016

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	33.98	0.04

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 2/29/2016
Certification Expiration: 2/28/2017
Simulator testing technician: R. WELSH

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: ELLEN STRAWSINE

Certification No: SD2252_2292016

X

DHSS BAP Scientist Approving