Missouri Department of Health and Senior Services
P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Peter Lyskowski
Director

Jeremiah W. (Jay) Nixon
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2285 **Manufacturer:** Guth

Model Number: 10-4D

Agency: GREENWOOD PD

Agency Address: 709 WEST MAIN ST., GREENWOOD, MO 64034

NIST THERMOMETER INFORMATION

Serial Number: 358440 **Bias:** 0.00

Uncertainty: 0.02

Date of Certification: 9/8/2015 **Date of Expiration:** 9/8/2016

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Simulator Average	NIST Average	Combined Uncertainty
34.00	34.00	0.03

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 1/14/2016
Certification Expiration: 1/14/2017
Simulator testing technician: D LUCAS

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: ELLEN STRAWSINE
Certification No: SD2285_1142016

X

DHSS BAP Scientist Approving