



Missouri Department of Health and Senior Services  
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



**Peter Lyskowski**  
 Director

**Jeremiah W. (Jay) Nixon**  
 Governor

# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** SD2727      **Manufacturer:** Guth  
**Model Number:** 10-4D  
**Agency:** LAMAR PD  
**Agency Address:** 1106 BROADWAY, LAMAR, MO 64759

## NIST THERMOMETER INFORMATION

**Serial Number:** 093752      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 9/8/2015      **Date of Expiration:** 9/8/2016

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	33.99	0.02

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 3/2/2016  
**Certification Expiration:** 3/2/2017  
**Simulator testing technician:** D. LUCAS

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** ELLEN STRAWSINE

**Certification No:** SD2727\_322016

X

DHSS BAP Scientist Approving