



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Peter Lyskowski
 Director

Jeremiah W. (Jay) Nixon
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2753 **Manufacturer:** Guth
Model Number: 10-4D
Agency: KIMBERLING CITY PD
Agency Address: PO BOX 370, KIMBERLING CITY, MO 65686

NIST THERMOMETER INFORMATION

Serial Number: 093752 **Bias:** 0.00
Uncertainty: 0.017
Date of Certification: 9/8/2015 **Date of Expiration:** 9/8/2016

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.00	0.018

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 1/26/2016
Certification Expiration: 1/26/2017
Simulator testing technician: D. DEBOARD

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIAN LUTMER
Certification No: SD2753_1262016

X 

DHSS BAP Scientist Approving