



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED  
 BY MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030794	PRINTER SN 84.9324.031	DATE OF INSPECTION 03/27/2020
-----------------------------	---------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 1781 Zumbehl Road, St. Charles, MO 63303	TIME OF INSPECTION 4:08 am
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
STANDARD SUPPLIER <u>Guth Laboratories</u> LOT # <u>19160</u> EXP. DATE <u>07/09/2021</u>	
SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0° C</u> SIMULATOR SN <u>SD2248</u> SIMULATOR EXP DATE <u>05/22/2020</u>	

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.097	TEST 2 ← .096	TEST 3 ← .097
----------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

New 9 Volt Battery Installed  
 Time Changed (Daylight Saving Time)

<b>INSPECTING OFFICER</b>	
SIGNATURE #440	PRINT NAME Stephen Gresham
TYPE II PERMIT NUMBER/EXPIRATION DATE 290245 10/16/2021	TELEPHONE NUMBER (636) 949-3300

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901



**GUTH LABORATORIES, INC.**

880 NORTH 67th STREET • HARRISBURG, PA 17111-4811 • TELEPHONE: 717-684-6470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19160** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**STEPHEN L GRESHAM**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/16/2019

NUMBER 290245

EXPIRES 10/16/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator GRESHAM, STEPHEN  
 Permit No 290245  
 Date issued 10/16/2019 Date Expires 10/16/2021



AS IV Serial no: 030794  
Version no: 0040

TEST RECORD 01852

Temp Date Time 210L

Air Blank:  
03/27/20 04:08 .000  
Subject Tests Auto  
22 03/27/20 04:08 .000

Subject Name

BLANK TEST

Subject I.D.

Operator Name, I.D.

GRESHAM 290245

Location

1781 ZUMBEHL ROAD

ST. CHARLES, MD 63303

AS IV Serial no: 030794  
Version no: 0040

TEST RECORD 01854

Temp Date Time 210L

Air Blank:  
03/27/20 04:14 .000  
Calibration Check:  
24 03/27/20 04:14 .097

Subject Name

TEST # 1

Subject I.D.

Operator Name, I.D.

GRESHAM 290245

Location

1781 ZUMBEHL ROAD

ST. CHARLES, MD 63303

AS IV Serial no: 030794  
Version no: 0040

TEST RECORD 01855

Temp Date Time 210L

Air Blank:  
03/27/20 04:16 .090  
Calibration Check:  
25 03/27/20 04:16 .096

Subject Name

TEST # 2

Subject I.D.

Operator Name, I.D.

GRESHAM 290245

Location

1781 ZUMBEHL ROAD

ST. CHARLES, MD 63303

AS IV Serial no: 030794  
Version no: 0040

TEST RECORD 01856

Temp Date Time 210L

Air Blank:  
03/27/20 04:12 .000  
Calibration Check:  
25 03/27/20 04:12 .097

Subject Name

TEST # 3

Subject I.D.

Operator Name, I.D.

GRESHAM 290245

Location

1781 ZUMBEHL ROAD

ST. CHARLES, MD 63303

AS IV Serial no: 030794  
Version no: 0040

TEST RECORD 01857

Temp Date Time 210L

Void: RFI  
12 03/27/20 04:19

Subject Name

RFI TEST

Subject I.D.

Operator Name, I.D.

GRESHAM 290245

Location

1781 ZUMBEHL ROAD

ST. CHARLES, MD 63303