

Airgas

Airgas USA LLC (Lab)
5775 Grand Street
St. Louis, Mo. 63103
Phone: 314-533-3100
Fax: 314-634-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2061 Craig Road
St. Louis, Mo 63146

Issue Date: 28-Aug-2018

Lot # AG824102 Model 106401

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Measured Concentration</u>
29-Aug-2020	108	Ethanol Nitrogen	218.5 (1002 BrAC (218 ppm))

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010583	393.0 ppm
EB0010570	289.8 ppm	EB0010585	288.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.8 ppm	EB0010562	104.2 ppm
EB0010584	52.12 ppm	EB0010573	52.91 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2018.08.28 18:38:29 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
SCOTT RONALD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/8/2020

NUMBER 200202

EXPIRES 7/8/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator RONALD, SCOTT
 Permit No 200202
 Date Issued 7/8/2020 Date Expires 7/8/2022

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00043

Temp Date Time ^{s/} 210L

Air Blank:
07/15/20 10:23 .000
Calibration Check:
28 07/15/20 10:23 .078

Subject Name

TEST

Subject I.D.

#1

Operator Name, I.D.

SCOTT RONALD, 200202

Location

SCCPD

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00044

Temp Date Time ^{s/} 210L

Air Blank:
07/15/20 10:24 .000
Calibration Check:
28 07/15/20 10:24 .078

Subject Name

TEST

Subject I.D.

#2

Operator Name, I.D.

SCOTT RONALD, 200202

Location

SCCPD

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00045

Temp Date Time ^{s/} 210L

Air Blank:
07/15/20 10:26 .000
Calibration Check:
28 07/15/20 10:26 .078

Subject Name

TEST

Subject I.D.

#3

Operator Name, I.D.

SCOTT RONALD, 200202

Location

SCCPD

AS IV Serial no: 034874
Version no: 532C

TEST RECORD 00046

Temp Date Time ^{s/} 210L

VOID: RFI
12 07/15/20 10:27

Subject Name

TEST

Subject I.D.

RFI

Operator Name, I.D.

SCOTT RONALD, 200202

Location

SCCPD