



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 BY: [unclear] DATE: 2/10/2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 043580	PRINTER SN 84.9324.045	DATE OF INSPECTION 02/10/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 101 Sheriff Dierker Ct, O'Fallon MO 63366		TIME OF INSPECTION 11:13 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)	Passed
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)	22°C
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY	Passed
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY	11:13 hours 02/10/2020

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters, Inc.</u>	LOT # <u>AG824102</u> EXP. DATE <u>08/29/2020</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____	SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input checked="" type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .077	TEST 2 .077	TEST 3 .077
--------------	--------------	--------------

RFI DETECTOR OPERATING Passed

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	(OVER .19)	3
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Officer Ginnever, DSN 620
TYPE II PERMIT NUMBER/EXPIRATION DATE 290152-07/17/2021	TELEPHONE NUMBER (636) 949-3000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

ST. CHARLES COUNTY POLICE DEPARTMENT - BREATH ALCOHOL PROGRAM

<p>AS IV Serial no: 043580 Version no: 532C</p> <p>TEST RECORD 00179</p> <p>Temp Date Time 210L ^{s/}</p> <p>Air Blank: 02/10/20 11:13 .000 Calibration Check: 22 02/10/20 11:13 .077</p> <p>Subject Name <u>TEST</u> Subject I.D. <u>Ginnem 620</u> Operator Name, I.D. <u>Ginnem 620</u> Location <u>SCDDOC</u></p>	<p>AS IV Serial no: 043580 Version no: 532C</p> <p>TEST RECORD 00180</p> <p>Temp Date Time 210L ^{s/}</p> <p>Air Blank: 02/10/20 11:14 .000 Calibration Check: 22 02/10/20 11:14 .077</p> <p>Subject Name <u>TEST</u> Subject I.D. <u>Ginnem 620</u> Operator Name, I.D. <u>Ginnem 620</u> Location <u>SCDDOC</u></p>	<p>AS IV Serial no: 043580 Version no: 532C</p> <p>TEST RECORD 00181</p> <p>Temp Date Time 210L ^{s/}</p> <p>Air Blank: 02/10/20 11:16 .000 Calibration Check: 22 02/10/20 11:16 .077</p> <p>Subject Name <u>TEST</u> Subject I.D. <u>Ginnem 620</u> Operator Name, I.D. <u>Ginnem 620</u> Location <u>SCDDOC</u></p>	<p>AS IV Serial no: 043580 Version no: 532C</p> <p>TEST RECORD 00182</p> <p>Temp Date Time 210L ^{s/}</p> <p>UO1D: RFI 12 02/10/20 11:17</p> <p>Subject Name <u>TEST</u> Subject I.D. <u>Ginnem 620</u> Operator Name, I.D. <u>Ginnem 620</u> Location <u>SCDDOC</u></p>
---	---	---	--

ASIV S/N: 043580
TYPE II: Ginnem 620

DATE: 02/10/2020

TICKET #1

TICKET #2

TICKET #3

TICKET #4



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 29-Aug-2018

Lot # AG824102 Model 108cacc

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
29-Aug-2020	108	Ethanol	0.080 ± 0.002 BrAC (218 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.5 ppm	EB0010562	104.2 ppm
EB0010584	52.12 ppm	EB0010579	52.81 ppm

Analytical Method: NDIR


Digitally signed by Quality Control
 Date: 2018.08.29 18:38:20 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3012.06



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

SCOTT R GINNEVER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 806.111 through 806.119 RSMo.

DATE 7/17/2019

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 290152

EXPIRES 7/17/2021

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-3, (PG-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GINNEVER, SCOTT
 Permit No 290152
 Date Issued 7/17/2019 Date Expires 7/17/2021