



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Tracy Crews at 7:52 am, Jul 09, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 062086	PRINTER SN 84.9324.048	DATE OF INSPECTION 06/09/2020
-----------------------------	---------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) # 1 Bruns Lane, Union MO 63084	TIME OF INSPECTION 6:48 am
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG902302 EXP. DATE 01/23/2021
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .101	TEST 2 ← .100	TEST 3 ← .100
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE <i>Deputy B. Erismann #1223</i>	PRINT NAME Deputy B. Erismann #1223
TYPE II PERMIT NUMBER/EXPIRATION DATE 200166 05/05/2022	TELEPHONE NUMBER (636) 583-2560

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 062086  
Version no: 532B

TEST RECORD 01020

Temp Date Time 210L

Air Blank:  
06/09/20 06:48 .000  
Calibration Check:  
21 06/09/20 06:48 .101

Subject Name

Test

Subject I.D.

Operator Name, I.D.

Erisman 1223

Location

AS IV Serial no: 062086  
Version no: 532B

TEST RECORD 01021

Temp Date Time 210L

Air Blank:  
06/09/20 06:50 .000  
Calibration Check:  
22 06/09/20 06:50 .100

Subject Name

Test

Subject I.D.

Operator Name, I.D.

Erisman 1223

Location

ASIV  
062086

AS IV Serial no: 062086  
Version no: 532B

TEST RECORD 01022

Temp Date Time 210L

Air Blank:  
06/09/20 06:53 .000  
Calibration Check:  
23 06/09/20 06:53 .100

Subject Name

Test

Subject I.D.

Operator Name, I.D.

Erisman 1223

Location

AS IV Serial no: 062086  
Version no: 532B

TEST RECORD 01023

Temp Date Time 210L

VOID: RFI  
12 06/09/20 06:57

Subject Name

Test

Subject I.D.

Operator Name, I.D.

Erisman 1223

Location



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 24-Jan-2019

**Lot # AG902302 Model 108cacd**

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
23-Jan-2021	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
 Date: 2019.01.24 10:46:58 -06:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

*Rod Marsala*  
 \_\_\_\_\_  
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**BRANDON ERISMAN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119, RSMo.

DATE 5/5/2020  
 NUMBER 200166  
 EXPIRES 5/5/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY  
  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (88-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator ERISMAN, BRANDON  
 Permit No 200166  
 Date Issued 5/5/2020 Date Expires 5/5/2022

STATE OF MISSOURI    )  
                                  )  
COUNTY OF FRANKLIN )       SS

**AFFIDAVIT FOR RECORDS**

*Before me, the undersigned authority personally appeared, Deputy B. Erisman #1223, and upon being duly sworn by me, deposed as follows:*

My name is Deputy B. Erisman #1223. I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am the custodian of the records of Franklin County Sheriff's Office, Alco Sensor IV serial# 062086. Attached hereto are 4 pages of records from the Franklin County Sheriff's Office for the date of June 9, 2020. These 4 pages of records are kept by the Franklin County Sheriff's Office in the regular course of business, and it is with the regular course of business that an employee or representative to make the record or to transmit information thereof to be included in such record. The records attached hereto are the original or exact duplicates of the original. The records attached hereto represent a complete and exact duplication of any and all original records kept by the Franklin County Sheriff's Office in the regular course of business.

Deputy B. Erisman #1223  
Affiant's Name – typed or printed

Dep B. Erisman #1223  
Affiant's Signature

*In witness whereof, I have hereunto subscribed my name and affixed my official seal this*  
9<sup>th</sup> day of June, 2020.

*My commission expires:* 09/14/2023

Kimberly A. Moritz  
Notary Public

