



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 9:19 am, Jul 06, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 062093	PRINTER SN 093.3563.014	DATE OF INSPECTION 07/02/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 1001 Heroes Way, Sugar Creek, MO. (Sugar Creek PD)	TIME OF INSPECTION 2230
--	----------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 19160 EXP. DATE 07/09/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD 3136 SIMULATOR EXP DATE 07/23/2020

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .102

TEST 2 ← .103

TEST 3 ← .105

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
 Garrett Holman

TYPE II PERMIT NUMBER/EXPIRATION DATE
 200126 03/09/2022

TELEPHONE NUMBER
 (816) 252-5560

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

SUGAR CREEK MISSOURI POLICE DEPARTMENT

ALCO-SENSOR IV WITH PRINTER TEST RESULTS

TEST # 1

IV Serial no: 062093
Version no: 532B

TEST RECORD 01241

Time Date Time 210L

Air Blank:
07/02/20 22:29 .000
Calibration Check:
22 07/02/20 22:29 .182

Subject Name

Holman

Subject I.D.

Test #1

Operator Name, I.D.

Location

SCPD

TEST # 2

IV Serial no: 062093
Version no: 532B

TEST RECORD 01242

Time Date Time 210L

Air Blank:
07/02/20 22:31 .000
Calibration Check:
23 07/02/20 22:31 .183

Subject Name

Holman

Subject I.D.

Test #2

Operator Name, I.D.

Location

SCPD

TEST # 3

IV Serial no: 062093
Version no: 532B

TEST RECORD 01243

Time Date Time 210L

Air Blank:
07/02/20 22:33 .000
Calibration Check:
23 07/02/20 22:33 .185

Subject Name

Holman

Subject I.D.

Test #3

Operator Name, I.D.

Location

SCPD

RFI TEST

IV Serial no: 062093
Version no: 532B

TEST RECORD 01244

Time Date Time 210L

RFI:
12 07/02/20 22:35

Subject Name

Holman

Subject I.D.

RFI TEST

Operator Name, I.D.

Location

SCPD



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 19160 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 10, 2019, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is July 9, 2021 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

GARRETT HOLMAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 506.111 through 506.119 RSMo.

DATE 3/9/2020

NUMBER 200126

EXPIRES 3/9/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 520-0771 (6-10)

LABORATION



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: HOLMAN, GARRETT
Permit No: 200126
Date Issued: 3/9/2020 Date Expires: 3/9/2022

