

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00705

Temp Date Time 210L

Air Blank:
01/17/20 00:22 .000
Calibration Check:
22 01/17/20 00:22 .081

Subject Name
Test

Subject I.D.
211

Operator Name, I.D.
P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00706

Temp Date Time 210L

Air Blank:
01/17/20 00:24 .000
Calibration Check:
23 01/17/20 00:24 .080

Subject Name
Test

Subject I.D.
211

Operator Name, I.D.
P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00707

Temp Date Time 210L

Air Blank:
01/17/20 00:26 .000
Calibration Check:
23 01/17/20 00:26 .081

Subject Name
Test

Subject I.D.
211

Operator Name, I.D.
P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00708

Temp Date Time 210L

VOID: RFI
12 01/17/20 00:27

Subject Name
Test

Subject I.D.
211

Operator Name, I.D.
P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00709

Temp Date Time 210L

Air Blank:
01/17/20 00:28 .000
Subject Test: Auto
24 01/17/20 00:28 .000

Subject Name
Test

Subject I.D.
211

Operator Name, I.D.
P.H.P.D.

Location



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 16-Jul-2019

Lot # AG919602 Model 108cadd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
15-Jul-2021	108	Ethanol	0.080 ± 0.002 BrAC (218 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.07.17 11:00:19 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
TODD D INSERRA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/4/2018

NUMBER 280288

EXPIRES 10/4/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator INSERRA, TODD
 Permit No 280288
 Date Issued 10/4/2018 Date Expires 10/4/2020

