



**MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 070763	PRINTER SN 083C.3556.249	DATE OF INSPECTION 03/23/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 8 Municipal, Park Hills, MO.		TIME OF INSPECTION 7:11 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeter INC. LOT # AG919602 EXP. DATE 07/15/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) N/A SIMULATOR SN N/A SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .079 TEST 2 ➔ .081 TEST 3 ➔ .083

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument is operating properly

INSPECTING OFFICER

SIGNATURE

PRINT NAME
Todd Inerra

TYPE II PERMIT NUMBER/EXPIRATION DATE
280288

TELEPHONE NUMBER
(573) 431-3122

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 070763
Version no: 532B
001101110: 0000

TEST RECORD 00718

Temp Date Time 210L

Air Blank:
03/23/20 07:13 .000
Calibration Check:
23 03/23/20 07:13 .081

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00717

Temp Date Time 210L

Air Blank:
03/23/20 07:11 .000
Calibration Check:
22 03/23/20 07:11 .079

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00719

Temp Date Time 210L

Air Blank:
03/23/20 07:14 .000
Calibration Check:
23 03/23/20 07:14 .083

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00720

Temp Date Time 210L

VOID: RFI
12 03/23/20 07:16

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00721

Temp Date Time 210L

Air Blank:
03/23/20 07:17 .000
Subject Test: Auto
24 03/23/20 07:17 .000

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 16-Jul-2019

Lot # AG919602 **Model** 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
15-Jul-2021	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (218 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
TODD D INSERRA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/4/2018

NUMBER 280288

EXPIRES 10/4/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator INSERRA, TODD
Permit No 280288
Date Issued 10/4/2018 **Date Expires** 10/4/2020

