



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 7:41 am, Dec 09, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 070763	PRINTER SN 083C.3556.249	DATE OF INSPECTION 12/08/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 8 Municipal, Park Hills, MO.		TIME OF INSPECTION 11:57 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeter INC. LOT # AG919602 EXP. DATE 07/15/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) N/A SIMULATOR SN N/A SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .080	TEST 2 ← .079	TEST 3 ← .079
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument is operating properly

INSPECTING OFFICER	
SIGNATURE/ 	PRINT NAME Todd Inserra
TYPE 11 PERMIT NUMBER/EXPIRATION DATE 280288	TELEPHONE NUMBER (573) 431-3122

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00754

Temp Date Time 210L

Air Blank:
12/08/20 11:57 .000
Calibration Check:
22 12/08/20 11:57 .080

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00755

Temp Date Time 210L

Air Blank:
12/08/20 11:59 .000
Calibration Check:
23 12/08/20 11:59 .079

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00756

Temp Date Time 210L

Air Blank:
12/08/20 12:01 .000
Calibration Check:
23 12/08/20 12:01 .079

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00757

Temp Date Time 210L

VOID: RFI
12 12/08/20 12:02

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location

AS IV Serial no: 070763
Version no: 532B

TEST RECORD 00758

Temp Date Time 210L

Air Blank:
12/08/20 12:03 .000
Subject Test: Auto
24 12/08/20 12:03 .000

Subject Name

Test

Subject I.D.

211

Operator Name, I.D.

P.H.P.D.

Location



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 16-Jul-2019

Lot # AG919602 **Model** 108cacd

Exp. Date

15-Jul-2021

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.080 ± 0.002 BrAC (218 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method:

NDIR

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
TODD D INSERRA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/4/2018

NUMBER 280288

EXPIRES 10/4/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator INSERRA, TODD
Permit No 280288
Date Issued 10/4/2018 **Date Expires** 10/4/2020

