



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED  
01/15/2020 11:43 AM

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087964	PRINTER SN 08C3556197	DATE OF INSPECTION 01/15/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 2040 E Hunter Nevada Mo 64772		TIME OF INSPECTION 1:43 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER intoximeters LOT # AG827002 EXP. DATE 09/27/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .081

TEST 2 • .081

TEST 3 • .080

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	1	(.05-.09)	1	(.10-.14)	1	(.15-.19)	0	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

New printer ribbon and paper put it.

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME A. Stimson
TYPE II PERMIT NUMBER/EXPIRATION DATE #290101 05/03/2021	TELEPHONE NUMBER (417) 283-2400

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**ALEXANDER K STIMSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/3/2019

NUMBER 290101

EXPIRES 5/3/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** STIMSON, ALEXANDER  
**Permit No** 290101  
**Date Issued** 5/3/2019 **Date Expires** 5/3/2021





Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 2-Oct-2018

Lot # AG827002 Model 108cadd


<u>Exp. Date</u> 27-Sep-2020	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (208 ppm) Balance
---------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

Analytical Method: NDIR

Digitally signed by Quality Control  
 Date: 2018.10.03 10:02:04 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release:   
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

AS IV Serial no: 087964  
Version no: 532B

TEST RECORD 00493  
Temp Date Time 210L

VOID: RFI  
12 01/15/20 01:52

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 087964  
Version no: 532B

TEST RECORD 00492  
Temp Date Time 210L

Air Blank:  
01/15/20 01:50 .000

Calibration Check:  
19 01/15/20 01:50 .000

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 087964  
Version no: 532B

TEST RECORD 00491  
Temp Date Time 210L

Air Blank:  
01/15/20 01:49 .000

Calibration Check:  
18 01/15/20 01:49 .001

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 087964  
Version no: 532B

TEST RECORD 00490  
Temp Date Time 210L

Air Blank:  
01/15/20 01:47 .000

Calibration Check:  
17 01/15/20 01:47 .001

Subject Name

Subject I.D.

Operator Name, I.D.

Location

A. Stinson/2901

A. Stinson/2901

A. Stinson/2901

A. Stinson/2901

VCSO

VCSO

VCSO

VCSO