



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 11:48 am, Mar 16, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087970	PRINTER SN 08c.3527.187	DATE OF INSPECTION 03/12/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 9701 Marion Park Dr. KCMO 64137		TIME OF INSPECTION 7:21 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u>	LOT # <u>AG905605</u> EXP. DATE <u>02/25/2021</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input checked="" type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.080	TEST 2 ← 0.080	TEST 3 ← 0.079
----------------	----------------	----------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS standards and guidelines.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Douglas Davidson 5646
TYPE II PERMIT NUMBER/EXPIRATION DATE 290087/04/22/2021	TELEPHONE NUMBER (816) 234-5000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT
TYPE II

DOUGLAS DAVIDSON



Airgas

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Test Date: 26-Feb-2019

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Lot # AG905605 Model 108cacc

Exp. Date
25-Feb-2021

CYL Type
108

Component
Ethanol
Nitrogen

Certified Concentration
0.082 ± 0.002 B7AC (223 ppm)
Balance

[Signature]
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (98-10)

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):
ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000
for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2019

NUMBER 290087

EXPIRES 4/22/2021

MO 99-9771 (6-10)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The manual certifier is authorized to operate an evidential breath alcohol instrument of this designation or the alcoholic content in breath form of expired air in Missouri.

Operator: DAVIDSON, DOUGLAS
Permit No: 290087
Date Issued: 4/22/2019 Date Expires: 4/22/2021

Digitally signed by Quality Control
Date: 2019.02.26 14:34:46 -0800
Reason: I am the originator of this certification of analysis
Location: Algas USA LLC (Lab)

Approved for Release: *[Signature]*
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07

AS IV Serial no: 087970
Version no: 532B

TEST RECORD 00268

Temp Date Time ^{s/} 210L

Air Blank:
03/12/20 19:55 .000
Calibration Check:
25 03/12/20 19:55 .000

Subject Name
Test 1
Subject I.D.

Operator Name, I.D.

Location
Davidson 290087

AS IV Serial no: 087970
Version no: 532B

TEST RECORD 00270

Temp Date Time ^{s/} 210L

Air Blank:
03/12/20 19:59 .000
Calibration Check:
26 03/12/20 19:59 .079

Subject Name
Test 3
Subject I.D.

Operator Name, I.D.
Davidson 290087
Location

AS IV Serial no: 087970
Version no: 532B

TEST RECORD 00269

Temp Date Time ^{s/} 210L

Air Blank:
03/12/20 19:57 .000
Calibration Check:
25 03/12/20 19:57 .000

Subject Name
Test 2
Subject I.D.

Operator Name, I.D.
Davidson 290087
Location

AS IV Serial no: 087970
Version no: 532B

TEST RECORD 00271

Temp Date Time ^{s/} 210L

VOID: RFI
12 03/12/20 20:01

Subject Name
RFI Test
Subject I.D.

Operator Name, I.D.

Location
Davidson 290087

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
**BLOOD ALCOHOL TEST REPORT - ALCO-SENSOR IV
 WITH PRINTER**

FORM #8

SUBJECT'S NAME	DATE OF TEST
----------------	--------------

OPERATIONAL CHECKLIST: ALCO-SENSOR IV WITH PRINTER

ALCO-SENSOR SERIAL NO. 087970	PRINTER SERIAL NO. 08C.3527.187	LOCATION OF INSTRUMENT
---	---	------------------------

- 1. Examination of mouth conducted. If any substance is observed or indicated to be present, the substance observed or indicated must be removed prior to starting the 15 minute observation period.
- 2. Subject observed for at least 15 minutes by PO Davidson 5646. No smoking, oral intake or vomiting during this time; if vomiting occurs, start over with 15 minute observation period.
- 3. Make sure printer is connected to Alco-Sensor IV.
- 4. Turn printer on.
- 5. Insert mouthpiece into Alco-Sensor IV.
- 6. Observe temperature display, make sure temperature reading is between 10°C and 40°C.
- 7. When "BLNK" is displayed on Alco-Sensor IV, air blank is taken.
- 8. When "TEST" is displayed on Alco-Sensor IV, take subject breath sample.
- 9. When "SET" is displayed on Alco-Sensor IV, press SET button.
- 10. When printer has completed printing test result, tear off tape and fill in subject and officer information.
- 11. Press red button to eject mouthpiece.
- 12. Turn printer off.
- 13. Attach printout to this report.

CERTIFICATION BY OPERATOR	BAC
----------------------------------	-----

As set forth in the rules promulgated by the Department of Health and Senior Services related to the determination of blood alcohol by breath analysis, I certify that:

- 1. There was no deviation from the procedure approved by the department.
- 2. To the best of my knowledge the instrument was functioning properly.
- 3. I am authorized to operate the instrument.
- 4. No radio transmission occurred inside the room where and when this was being conducted.

NAME OF OPERATOR Davidson	PERMIT NO. 290087	EXPIRATION DATE 04/22/2021
WITNESS (IF ANY)		DATE

