

AS IV Serial no: 094801
Version no: 532B

TEST RECORD 00476

Temp Date Time 210L

Air Blank:
02/01/20 23:10 .000
Calibration Check:
16 02/01/20 23:10 .104

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

E. Stock 280121

Location

OAKVIEW PD

AS IV Serial no: 094801
Version no: 532B

TEST RECORD 00477

Temp Date Time 210L

Air Blank:
02/01/20 23:14 .000
Calibration Check:
18 02/01/20 23:14 .104

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

E. Stock 280121

Location

OAKVIEW PD

AS IV Serial no: 094801
Version no: 532B

TEST RECORD 00478

Temp Date Time 210L

Air Blank:
02/01/20 23:16 .000
Calibration Check:
19 02/01/20 23:16 .102

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

E. Stock 280121

Location

OAKVIEW PD

AS IV Serial no: 094801
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00479

Temp Date Time 210L

VOID: RFI
12 02/01/20 23:21

Subject Name

RFI TEST

Subject I.D.

Operator Name, I.D.

E. Stock 280121

Location

OAKVIEW PD

AS IV Serial no: 094801
Version no: 532B

TEST RECORD 00480

Temp Date Time 210L

Air Blank:
02/01/20 23:23 .000
Subject Test: Auto
23 02/01/20 23:23 .000

Subject Name

Blank TEST

Subject I.D.

Operator Name, I.D.

E. Stock 280121

Location

OAKVIEW PD



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 26-Mar-2019

Lot # AG908401 Model 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
25-Mar-2021	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.03.26 12:19:36 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: 
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07