



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 095961	PRINTER SN Nixa 099.3586.643	DATE OF INSPECTION 12/21/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 715 W Center Circle, Nixa MO 65714		TIME OF INSPECTION 23:00

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCo Marketing Inc LOT # 19002 EXP. DATE 10/16/2021
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN MP5537 SIMULATOR EXP DATE 09/16/2021
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100	TEST 2 .100	TEST 3 .099
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Meets or Exceeds Department of Health and Senior Services Rules and Regulations.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Sgt. R Seiner
TYPE II PERMIT NUMBER/EXPIRATION DATE 200216 07/30/2022	TELEPHONE NUMBER (417) 725-2510

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

Nixa Police Department

Calibration Check slip's

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02090

Temp	Date	Time	a/
	12/22/20	23:04	.000
	24 12/22/20	23:04	.100

Air Blank:
12/22/20 23:04 .000
Calibration Check:
24 12/22/20 23:04 .100

Subject Name
TEST 1

Subject I.D.

Operator Name, I.D.
RSEINER 200216

Location
NIXA PD

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02091

Temp	Date	Time	a/
	12/22/20	23:06	.000
	25 12/22/20	23:06	.100

Air Blank:
12/22/20 23:06 .000
Calibration Check:
25 12/22/20 23:06 .100

Subject Name
TEST 2

Subject I.D.

Operator Name, I.D.
RSEINER 200216

Location
NIXA PD

AS IV Serial no: 095961
Version no: 532B

TEST RECORD 02092

Temp	Date	Time	a/
	12/22/20	23:08	.000
	26 12/22/20	23:08	.000

Air Blank:
12/22/20 23:08 .000
Calibration Check:
26 12/22/20 23:08 .000

Subject Name
TEST 3

Subject I.D.

Operator Name, I.D.
RSEINER 200216

Location
NIXA PD

Nixa Police Department

Blank (Zero) test Evidence slip

AS DU Serial no: 095961
Version no: 582B

TEST RECORD: 02093

Temp Date Time 210L

Air Blank:
12/22/20 23:10 .000
Subject Test: Auto
26 12/22/20 23:10 .000

Subject Name

BLANK TEST

Subject I.D.

Operator Name, I.D.

RSENER 200216

Location

NIXA PD

Nixa Police Department

RFI Evidence slip

AS 10 Serial no: 095961
Version no: 532B

TEST RECORD 02094
Temp Date Time 210L

VOID: RFI
12/22/20 23:11

Subject Name

RFI TEST

Subject I.D.

Operator Name: I.D.

RSEINER 200216

Location

NIXA PD



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ROBERT A SEINER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/30/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 200216

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 7/30/2022

MO 680-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SEINER, ROBERT
 Permit No 200216
 Date Issued 7/30/2020 Date Expires 7/30/2022



STATE OF MISSOURI)
)
COUNTY OF CHRISTIAN)

AFFIDAVIT

Before me, the undersigned authority, personally appeared Robert Seiner, who, being by me duly sworn, deposed as follows:

My name is Robert Seiner, I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:


I am the custodian of records for the Nixa Police Department Alco-Sensor IV with printer, located at the Nixa Police Department. Attached hereto are 6 pages from my Alco-Sensor IV with printer records. The 6 pages of records are kept by me, in the regular course of business, and it was the regular course of business of the Nixa Police Department for an employee or representative of the Nixa Police Department, with the knowledge of the act, event, condition, opinion, or diagnosis recorded to make the record or to transmit information thereof to be included in such record; and the record was made at or near the time of the act, event, condition, opinion, or diagnosis. The records attached hereto are the original or exact duplicates of the original.

Affiant

In witness whereof, I have hereunto subscribed my name and affixed my official seal this ___ day of _____, 2020.

Notary Public, County of Christian

(seal)



RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 19002

EXPIRATION DATE: October 16, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer, Operations Manager
RepCo Marketing Co.