



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097403	PRINTER SN 099.3586.830	DATE OF INSPECTION 04/12/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 7447 Dale Avenue, Saint Louis		TIME OF INSPECTION 11:40

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG814202 EXP. DATE 05/22/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .082

TEST 2 ← .085

TEST 3 ← .083

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

NONE

INSPECTING OFFICER

SIGNATURE 	PRINT NAME CPL Richard Erwin
---------------	---------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 290013 01-17-2021	TELEPHONE NUMBER (314) 647-5656
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

1

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00304

Temp Date Time ^{9/} 210L

Air Blank:
05/08/20 20:53 .000
Calibration Check:
23 05/08/20 20:53 .079

Subject Name

Joe Doe

Subject I.D.

123

Operator Name, I.D.

Erin 167

Location

7447 Dale Ave

2

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00305

Temp Date Time ^{9/} 210L

Air Blank:
05/08/20 20:54 .000
Calibration Check:
24 05/08/20 20:54 .078

Subject Name

Joe Doe

Subject I.D.

123

Operator Name, I.D.

Erin 167

Location

7447 Dale Ave

3

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00306

Temp Date Time ^{9/} 210L

Air Blank:
05/08/20 20:56 .000
Calibration Check:
24 05/08/20 20:56 .079

Subject Name

Joe Doe

Subject I.D.

123

Operator Name, I.D.

Erin 167

Location

7447 Dale Ave

AS IV Serial no: 097403
Version no: 532B

TEST RECORD 00303

Temp Date Time ^{9/} 210L

~~VOID: REI~~
12 05/08/20 20:51

Subject Name

Joe Doe

Subject I.D.

123

Operator Name, I.D.

Erin 167

Location

7447 Dale Ave



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

RICHARD A ERWIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/17/2019

NUMBER 290013


EXPIRES 1/17/2021

MO 680-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)


 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ERWIN, RICHARD
Permit No 290013
Date issued 1/17/2019 Date Expires 1/17/2021





Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 23-May-2018

Lot # AG814202 Model 55cacd

<u>Exp. Date</u> 22-May-2020	<u>Cyl. Type</u> 55	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (218 ppm) Balance
---------------------------------	------------------------	---	---

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	269.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010585	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010679	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2018.05.24 13:46:42 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06