



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**

By Tracy Crews at 10:14 am, Sep 24, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>097409</u>	PRINTER SN <u>096.3580.871</u>	DATE OF INSPECTION <u>8-4-2020</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>602 S. Franklin Cuba</u>		TIME OF INSPECTION <u>2:08 am</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS-OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Cuth LOT # 20190 EXP. DATE 4-6-22
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34° SIMULATOR SN MP3872 SIMULATOR EXP DATE 5-7-2021

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <sup>100</sup> <u>.098</u>	TEST 2 <sup>100</sup> <u>.100</u>	TEST 3 <sup>100</sup> <u>.100</u>
-----------------------------------	-----------------------------------	-----------------------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>0</u>	(OVER .19) <u>0</u>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE <u>Michael Condorec</u>	PRINT NAME <u>Michael Condorec</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>200163/5-4-2022</u>	TELEPHONE NUMBER <u>573-885-7979</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IU Serial no: 097409  
Version no: 004C

TEST RECORD 00566

Temp Date Time 210L  
%/  
Air Blank: 08/04/20 02:08 .000  
Calibration Check: 21 08/04/20 02:08 .095

Subject Name

Blank

Subject I.D.

Test 1

Operator Name, I.D.

Castro / 200163

Location

602 S. Franklin

Cuba

AS IU Serial no: 097409  
Version no: 004C

TEST RECORD 00567

Temp Date Time 210L  
%/  
Air Blank: 08/04/20 02:10 .000  
Calibration Check: 21 08/04/20 02:10 .100

Subject Name

Blank

Subject I.D.

Test 2

Operator Name, I.D.

Castro / 200163

Location

602 S. Franklin

Cuba

AS IU Serial no: 097409  
Version no: 004C

TEST RECORD 00568

Temp Date Time 210L  
%/  
Air Blank: 08/04/20 02:12 .000  
Calibration Check: 22 08/04/20 02:12 .100

Subject Name

Blank

Subject I.D.

Test 3

Operator Name, I.D.

Castro / 200163

Location

602 S. Franklin

Cuba

AS IU Serial no: 097409  
Version no: 004C

TEST RECORD 00569

Temp Date Time 210L  
%/  
Void: NFI  
12 08/04/20 02:13

Subject Name

Blank

Subject I.D.

RPT

Operator Name, I.D.

Castro / 200163

Location

602 S. Franklin

Cuba



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 8, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 6, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L**  $\pm$  3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



PERMIT  
TYPE II

MICHAEL SCENTUNZI

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 506.111 through 506.119 RSMo.

DATE 5/4/2020

NUMBER 200163

EXPIRES 5/4/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 550-0771 (5-19)

LAB 4 (08-19)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named Controller is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator: SCENTUNZI, MICHAEL  
Permit No: 200163  
Date Issued: 5/4/2020 Date Expires: 5/4/2022

