



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
BY: [unclear] [unclear] at 8:02 am, Nov 13, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|--|----------------------------|----------------------------------|
| ALCO SENSOR IV SN 097411 | PRINTER SN 03A.2436.096 | DATE OF INSPECTION 11/11/2020 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 4001 NE Lakewood Way Ct., Lee's Summit, Mo 64064 | | TIME OF INSPECTION 9:36 pm |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG006306 EXP. DATE 03-03-2022
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|--------------|--------------|--------------|
| TEST 1 .102 | TEST 2 .102 | TEST 3 .101 |
|--------------|--------------|--------------|

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 2 | (0-.04) | 0 | (.05-.09) | 1 | (.10-.14) | 2 | (.15-.19) | 3 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DHSS standards and guidelines. Time change completed after maintenance

| | |
|--|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE | PRINT NAME S. Plain #101 |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 200054 01/10/2022 | TELEPHONE NUMBER (816) 795-1960 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

MS JU Serial no: 097411
Version no: 532B

TEST RECORD 00991
Temp Date Time 2100

Air Blank:
12/11/20 22:41

Calibration Check:
22/11/20 22:41

Subject Name
Mont. Maint.

Subject I.D.
-

Operator Name: I.D.
Rep. SPC #107

Location
JCSO 6HQ

MS JU Serial no: 097411
Version no: 532B

TEST RECORD 00990
Temp Date Time 2100

Air Blank:
11/11/20 22:40

Calibration Check:
22/11/20 22:40

Subject Name
Mont. Maint.

Subject I.D.
-

Operator Name: I.D.
Rep. SPC #107

Location
JCSO 6HQ

MS JU Serial no: 097411
Version no: 532B

TEST RECORD 00989
Temp Date Time 2100

Air Blank:
11/11/20 22:38

Calibration Check:
21/11/20 22:38

Subject Name
Mont. Maint.

Subject I.D.
-

Operator Name: I.D.
Rep. SPC #107

Location
JCSO 6HQ

MS JU Serial no: 097411
Version no: 532B

TEST RECORD 00988
Temp Date Time 2100

Air Blank:
11/11/20 22:36

Calibration Check:
20/11/20 22:36

Subject Name
Mont. Maint.

Subject I.D.
-

Operator Name: I.D.
Rep. Pdn #107

Location
JCSO 6HQ



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 4-Mar-2020

Lot # AG006306 Model 108cacc

Exp. Date

3-Mar-2022

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

RGM Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

CRM Serial No.

CC434668

CC234503

Concentration

800.0 ppm

253.0 ppm

CRM Serial No.

0056649

0056662

Concentration

390.1 ppm

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2020.03.05 13:27:24 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
SEAN PLAIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **1/10/2020**

NUMBER **200054**

EXPIRES **1/10/2022**

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-G771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PLAIN, SEAN
 Permit No 200054
 Date Issued 1/10/2020 Date Expires 1/10/2022