

RECEIVED

By Tracy Crews at 7:31 am, Sep 10, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097145 NAME OF AGENCY Missouri State Highway Patrol DATE OF INSPECTION 09/02/2020

LOCATION OF INSTRUMENT (STREET AND CITY) Troop D, Zone 3, Zone Office, Bolivar TIME OF INSPECTION 11:28

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

[X] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

[X] TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 21°C

[X] PRINTER WORKING PROPERLY

[X] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

[X] SIMULATOR SOLUTION [ ] COMPRESSED ETHANOL-GAS MIXTURE

[X] STANDARD SUPPLIER RepCo Marketing LOT # 19001 EXP. DATE 03-12-21

[X] SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP 2146 SIM. NIST EXP DATE 01-16-2021

[X] CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [X] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
[ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
[ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 098 TEST 2 ← .096 TEST 3 ← .096

[X] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

REPLACE PRINTER PAPER

INSPECTING OFFICER

SIGNATURE [Signature]

PRINT NAME Joshua White

TYPE II PERMIT NUMBER/EXPIRATION DATE 200042/ 01-07-2022

TELEPHONE NUMBER (417 ) 895-6868

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

TEST RECORD - REPRINT

TEST RECORD 00700

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/02/20 11:28 .000  
Calibration Check:  
21 09/02/20 11:28 .098

Subject Name

Subject I.D.

Operator Name, I.D.

TPR J.L. WHITE 200042

Location

TROOP D ZONE 3  
ZONE OFFICE  
BOLZUAR

TEST RECORD 00701

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/02/20 11:35 .000  
Calibration Check:  
23 09/02/20 11:35 .096

Subject Name

Subject I.D.

Operator Name, I.D.

TPR J.L. WHITE 200042

Location

TROOP D ZONE 3  
ZONE OFFICE  
BOLZUAR

TEST RECORD 00702

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/02/20 11:36 .000  
Calibration Check:  
24 09/02/20 11:36 .096

Subject Name

Subject I.D.

Operator Name, I.D.

TPR J.L. WHITE 200042

Location

TROOP D ZONE 3  
ZONE OFFICE  
BOLZUAR

TEST RECORD 00703

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 09/02/20 11:37

Subject Name

Subject I.D.

Operator Name, I.D.

TPR J.L. WHITE 200042

Location

TROOP D ZONE 3  
ZONE OFFICE  
BOLZUAR

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.**

**LOT NUMBER: 19001**

**EXPIRATION DATE: March 12, 2021 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1210 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 13, 2019  
The expiration date for this lot number is March 12, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

  
Alma Palmer, Operations Manager  
RepCo Marketing Co.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II  
JOSHUA L WHITE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/7/2020

NUMBER 200042

EXPIRES 1/7/2022

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **WHITE, JOSHUA**  
Permit No **200042**  
Date Issued **1/7/2020**    Date Expires **1/7/2022**