



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

RECEIVED

By Tracy Crews at 9:21 am, Jan 02, 2020

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097419	PRINTER SN 096.3580.861	DATE OF INSPECTION 01/01/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 200 N Main St Palmyra	TIME OF INSPECTION 9:13 am
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u>	LOT # <u>18370</u> EXP. DATE <u>12/05/2020</u>
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.0</u>	SIMULATOR SN <u>DR5308</u> SIMULATOR EXP DATE <u>07/08/2020</u>

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .102	TEST 2 • .101	TEST 3 • .102
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>Ronald C. Peer, Jr.</i>	PRINT NAME Ronald C. Peer, Jr.
TYPE II PERMIT NUMBER/EXPIRATION DATE 280224/07-03-2020	TELEPHONE NUMBER (573) 769-5540

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097419
Version no: 532B

TEST RECORD 00549

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
01/01/20 09:13 .000
Calibration Check:
19 01/01/20 09:13 .102

Subject Name

CALIBRATION CHECK
Subject I.D.

Operator Name, I.D.

Ronald C Bay #501
Location
PALMYRA PD

AS IV Serial no: 097419
Version no: 532B

TEST RECORD 00550

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
01/01/20 09:15 .000
Calibration Check:
20 01/01/20 09:15 .101

Subject Name

CALIBRATION CHECK
Subject I.D.

Operator Name, I.D.

Ronald C Bay #501
Location
PALMYRA PD

AS IV Serial no: 097419
Version no: 532B

TEST RECORD 00551

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
01/01/20 09:17 .000
Calibration Check:
21 01/01/20 09:17 .102

Subject Name

CALIBRATION CHECK
Subject I.D.

Operator Name, I.D.

Ronald C Bay #501
Location
PALMYRA PD

AS IV Serial no: 097419
Version no: 532B

TEST RECORD 00552

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 01/01/20 09:18

Subject Name

RFI TEST
Subject I.D.

Operator Name, I.D.

Ronald C Bay #501
Location
PALMYRA PD



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **18370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 6, 2018**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1218%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 5, 2020** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
RONALD C PEER JR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/3/2018

NUMBER 280224


EXPIRES 7/3/2020

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES


LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator PEER JR, RONALD
Permit No 280224
Date Issued 7/3/2018 Date Expires 7/3/2020