



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097422	PRINTER SN 097.3584.042	DATE OF INSPECTION 05/26/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 517 ONEIDA ST. SENECA, MISSOURI	TIME OF INSPECTION 9:30 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
--	---

<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u> LOT # <u>19370</u> EXP. DATE <u>12/09/2021</u>

<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>33.99</u> SIMULATOR SN <u>MP 2851</u> SIMULATOR EXP DATE <u>03/03/2021</u>
--

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➔ .103	TEST 2 ➔ .102	TEST 3 ➔ .102
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Joshua Fort
---------------	---------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 200125 03/09/2022	TELEPHONE NUMBER (417) 776-8158
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097422
Version no: 532B

TEST RECORD 00344

Temp Date Time ^{s/} 210L

Air Blank:
05/26/20 21:30 .000
Calibration Check:
22 05/26/20 21:30 .103

Subject Name

Test

Subject I.D.

1

Operator Name, I.D.

Joshua Fort 200125

Location Exp 3/9/22

517 Oneida St

Seneca, Mo 64865

AS IV Serial no: 097422
Version no: 532B

TEST RECORD 00345

Temp Date Time ^{s/} 210L

Air Blank:
05/26/20 21:31 .000
Calibration Check:
22 05/26/20 21:31 .102

Subject Name

Test

Subject I.D.

2

Operator Name, I.D.

Joshua Fort 200125

Location Exp 3/9/22

517 Oneida St

Seneca, Mo 64865

AS IV Serial no: 097422
Version no: 532B

TEST RECORD 00346

Temp Date Time ^{s/} 210L

Air Blank:
05/26/20 21:33 .000
Calibration Check:
23 05/26/20 21:33 .102

Subject Name

Test

Subject I.D.

3

Operator Name, I.D.

Joshua Fort 200125

Location Exp 3/9/22

517 Oneida St

Seneca, Mo 64865

AS IV Serial no: 097422
Version no: 532B

TEST RECORD 00347

Temp Date Time ^{s/} 210L

VOID: RFI
12 05/26/20 21:35

Subject Name

Test

Subject I.D.

RFI

Operator Name, I.D.

Joshua Fort 200125

Location Exp 3/9/22

517 Oneida St

Seneca, Mo 64865

AS IV Serial no: 097422
Version no: 532B

TEST RECORD 00348

Temp Date Time ^{s/} 210L

Air Blank:
05/26/20 21:36 .000
Subject Test: Auto
24 05/26/20 21:36 .000

Subject Name

Test

Subject I.D.

Blank

Operator Name, I.D.

Joshua Fort 200125

Location Exp 3/9/22

517 Oneida St

Seneca, Mo 64865



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JOSHUA FORT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/9/2020

NUMBER 200125

EXPIRES 3/9/2022

MO:580-0771 (6-10)


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator FORT, JOSHUA
Permit No 200125
Date issued 3/9/2020 **Date Expires** 3/9/2022





GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1199%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.