



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097422	PRINTER SN 097.3584.042	DATE OF INSPECTION 08/02/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 517 ONEIDA ST. SENECA, MISSOURI	TIME OF INSPECTION 9:52 pm
---	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
--	---

<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth Laboratories</u> LOT # <u>19370</u> EXP. DATE <u>12/09/2021</u>
---

<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>33.99</u> SIMULATOR SN <u>MP 2851</u> SIMULATOR EXP DATE <u>03/03/2021</u>
--

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .104	TEST 2  .103	TEST 3  .103
--------------	--------------	--------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME Joshua Fort
---------------	---------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 200125 03/09/2022	TELEPHONE NUMBER (417) 776-8158
--	------------------------------------

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS IV Serial no: 097422  
Version no: 532B

TEST RECORD 00356

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
08/02/20 21:52 .000  
Calibration Check:  
21 08/02/20 21:52 .104

Subject Name

Test

Subject I.D.

# 1

Operator Name, I.D.

Joshua Fort 200125

Location Exp. 3/9/22

517 Oneida St

Seneca, Mo 64865

AS IV Serial no: 097422  
Version no: 532B

TEST RECORD 00357

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
08/02/20 21:54 .000  
Calibration Check:  
22 08/02/20 21:54 .103

Subject Name

Test

Subject I.D.

# 2

Operator Name, I.D.

Joshua Fort 200125

Location Exp. 3/9/22

517 Oneida St

Seneca, Mo 64865

AS IV Serial no: 097422  
Version no: 532B

TEST RECORD 00358

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
08/02/20 21:57 .000  
Calibration Check:  
22 08/02/20 21:57 .103

Subject Name

Test

Subject I.D.

# 3

Operator Name, I.D.

Joshua Fort 200125

Location Exp. 3/9/22

517 Oneida St

Seneca, Mo 64865

AS IV Serial no: 097422  
Version no: 532B

TEST RECORD 00359

Temp Date Time <sup>a/</sup> 210L

VOID: RFI  
12 08/02/20 21:58

Subject Name

Test

Subject I.D.

RFI

Operator Name, I.D.

Joshua Fort 200125

Location Exp. 3/9/22

517 Oneida St

Seneca, Mo 64865

AS IV Serial no: 097422  
Version no: 532B

TEST RECORD 00360

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
08/02/20 21:59 .000  
Subject Test: Auto  
23 08/02/20 21:59 .000

Subject Name

Test

Subject I.D.

Blank

Operator Name, I.D.

Joshua Fort 200125

Location Exp. 3/9/22

517 Oneida St

Seneca, Mo 64865



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1199%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JOSHUA FORT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/9/2020

NUMBER 200125

EXPIRES 3/9/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (16-10)

MO-880-0771 (0-10)

**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator **FORT, JOSHUA**  
Permit No **200125**  
Date Issued **3/9/2020** Date Expires **3/9/2022**