

AS IV Serial no: 097425
Version no: 532B

TEST RECORD 00407

Temp Date Time ^{s/} 210L

Air Blank:
06/10/20 13:40 .000
Calibration Check:
21 06/10/20 13:40 .098

Subject Name

Cal Check #1

Subject I.D.

N/A

Operator Name, I.D.

Jahns 3725

Location

MoDot TMC

AS IV Serial no: 097425
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00408

Temp Date Time ^{s/} 210L

Air Blank:
06/10/20 13:42 .000
Calibration Check:
22 06/10/20 13:42 .098

Subject Name

Cal Check #2

Subject I.D.

N/A

Operator Name, I.D.

Jahns 3725

Location

MoDot TMC

AS IV Serial no: 097425
Version no: 532B

TEST RECORD 00409

Temp Date Time ^{s/} 210L

Air Blank:
06/10/20 13:46 .000
Calibration Check:
23 06/10/20 13:46 .098

Subject Name

Cal Check #3

Subject I.D.

N/A

Operator Name, I.D.

Jahns 3725

Location

MoDot TMC

AS IV Serial no: 097425
Version no: 532B

TEST RECORD 00410

Temp Date Time ^{s/} 210L

VOID: RFI
12 06/10/20 13:47

Subject Name

RFI Check

Subject I.D.

N/A

Operator Name, I.D.

Jahns 3725

Location

MoDot TMC



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 25-Oct-2018

Lot # AG829708 Model 108cadd

Exp. Date

24-Oct-2020

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)

Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2018.10.25 14:13:39 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**


KYLE JAHNS

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2019


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 290039

EXPIRES 2/20/2021


 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES