



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097440	PRINTER SN 096.3580.989	DATE OF INSPECTION 06/23/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Zone D-14 Office / Cassville, MO	TIME OF INSPECTION 3:55 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing, Inc. LOT # 19001 EXP. DATE 03/12/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIMULATOR SN MP2467 SIMULATOR EXP DATE 01/07/2021

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.102%

TEST 2 ← 0.100%

TEST 3 ← 0.100%

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibration performed.

INSPECTING OFFICER

SIGNATURE *D.R. Hukill #383*

PRINT NAME
Tpr. D.R. Hukill / #383

TYPE II PERMIT NUMBER/EXPIRATION DATE
290043 / 02/26/2021

TELEPHONE NUMBER
(417) 895-6868

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097440
Version no: 532B

TEST RECORD 00342

Temp Date Time 210L
s/

Air Blank:

06/23/20 16:03 .000

Calibration Check:
27 06/23/20 16:03 .102

Subject Name

Maint.

Subject I.D.

Test # 1

Operator Name, I.D.

Mr. D. Hukill #383

Location

Zone D-14

Cassville, MO

AS IV Serial no: 097440
Version no: 532B

TEST RECORD 00343

Temp Date Time 210L
s/

Air Blank:

06/23/20 16:05 .000

Calibration Check:
27 06/23/20 16:05 .100

Subject Name

Maint

Subject I.D.

Test # 2

Operator Name, I.D.

Mr. D. Hukill #383

Location

Zone D-14

Cassville, MO

AS IV Serial no: 097440
Version no: 532B

TEST RECORD 00344

Temp Date Time 210L
s/

Air Blank:

06/23/20 16:07 .000

Calibration Check:
28 06/23/20 16:07 .100

Subject Name

Maint

Subject I.D.

Test # 3

Operator Name, I.D.

Mr. D. Hukill #383

Location

Zone D-14

Cassville, MO

AS IV Serial no: 097440
Version no: 532B

TEST RECORD 00345

Temp Date Time 210L
s/

VOID: RFI

12 06/23/20 16:08

Subject Name

Maint.

Subject I.D.

RFI Test

Operator Name, I.D.

Mr. D. Hukill #383

Location

Zone D-14

Cassville, MO

Subject Name
Maint.
Subject I.D.
Calibration
Operator Name, I.D.
Mr. D. Hukill #383
Location
Zone D-14
Cassville, MO

Air Blank:
06/23/20 15:58 .000
Calibration:
25 06/23/20 15:58 .100

Temp Date Time 210L
s/

TEST RECORD 00341
Version no: 532B
AS IV Serial no: 097440

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 19001
EXPIRATION DATE: March 12, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1210 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 13, 2019
The expiration date for this lot number is March 12, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
DARWIN R HUKILL

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/26/2019

NUMBER 290043

EXPIRES 2/26/2021

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HUKILL, DARWIN
 Permit No 290043
 Date Issued 2/26/2019 Date Expires 2/26/2021