

A

AS IV Serial no: 097443
Version no: 532B

TEST RECORD 00961

Temp Date Time 210L
s/

Air Blank:
01/08/20 17:34 .000
Calibration Check:
24 01/08/20 17:34 .000

Subject Name

Self Test

Subject I.D.

Monthly Maint

Operator Name, I.D.

K Stults 290292

Location

Stone Co Jail

A

AS IV Serial no: 097443
Version no: 532B

TEST RECORD 00962

Temp Date Time 210L
s/

Air Blank:
01/08/20 17:37 .000
Calibration Check:
25 01/08/20 17:37 .098

Subject Name

Test 1

Subject I.D.

Monthly Maint

Operator Name, I.D.

K Stults 290292

Location

Stone Co Jail

A

AS IV Serial no: 097443
Version no: 532B

TEST RECORD 00963

Temp Date Time 210L
s/

Air Blank:
01/08/20 17:39 .000
Calibration Check:
26 01/08/20 17:39 .098

Subject Name

Test 2

Subject I.D.

Monthly Maint

Operator Name, I.D.

K Stults 290292

Location

Stone Co Jail

A

AS IV Serial no: 097443
Version no: 532B

TEST RECORD 00964

Temp Date Time 210L
s/

Air Blank:
01/08/20 17:43 .000
Calibration Check:
26 01/08/20 17:43 .097

Subject Name

Test 3

Subject I.D.

Monthly Maint

Operator Name, I.D.

K Stults 290292

Location

A

AS IV Serial no: 097443
Version no: 532B

TEST RECORD 00965

Temp Date Time 210L
s/

VOID: RFI
12 01/08/20 17:45

Subject Name

RFI

Subject I.D.

Monthly Maint

Operator Name, I.D.

K Stults 290292

Location

Stone Co Jail

Monthly Maint
01/08/2020

Stone Co. Sheriff's office

Dep. Kyle Stults

Permit # 290292

Exp: 12/16/2021



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 19160 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 10, 2019, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is July 9, 2021 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KYLE G STULTS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/16/2019

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 290292

EXPIRES 12/16/2021

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator STULTS, KYLE
 Permit No 290292
 Date Issued 12/16/2019 Date Expires 12/16/2021