

RECEIVED

By Stephen Wilson at 1:59 pm, May 12, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097445	PRINTER SN 096.3580.951	DATE OF INSPECTION 05/12/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 403 Highway 134 Kaiser, MO 65047		TIME OF INSPECTION 12:35 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth LOT # 18370 EXP. DATE 12/05/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIMULATOR SN MP4942 SIMULATOR EXP DATE 06/04/2020

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 \bullet .102

TEST 2 \bullet .102

TEST 3 \bullet .102

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Moved time up one hour.

INSPECTING OFFICER

SIGNATURE *Tyler S. Roby*

TYPE II PERMIT NUMBER/EXPIRATION DATE
200057; 01/10/2022

PRINT NAME
Tyler S. Roby

TELEPHONE NUMBER
(573) 508-2311

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097445
Version no: 532B

TEST RECORD 00551

Temp Date Time ^{s/} 210L

Air Blank:
05/12/20 12:43 .000
Calibration Check:
20 05/12/20 12:43 .102

Subject Name

Test

Subject I.D.

Test 1

Operator Name, I.D.

Tyler Roby, 200057

Location

Lake Oaark State Park

AS IV Serial no: 097445
Version no: 532B

TEST RECORD 00552

Temp Date Time ^{s/} 210L

Air Blank:
05/12/20 12:45 .000
Calibration Check:
21 05/12/20 12:45 .102

Subject Name

Test

Subject I.D.

Test 2

Operator Name, I.D.

Tyler Roby, 200057

Location

Lake Oaark State Park

AS IV Serial no: 097445
Version no: 532B

TEST RECORD 00553

Temp Date Time ^{s/} 210L

Air Blank:
05/12/20 12:46 .000
Calibration Check:
21 05/12/20 12:46 .102

Subject Name

Test

Subject I.D.

Test 3

Operator Name, I.D.

Tyler Roby, 200057

Location

Lake Oaark State Park

AS IV Serial no: 097445
Version no: 532B

TEST RECORD 00554

Temp Date Time ^{s/} 210L

VOID: RFI
12 05/12/20 12:48

Subject Name

Test - RFI

Subject I.D.

Test 4 - RFI

Operator Name, I.D.

Tyler Roby, 200057

Location

Lake Oaark State Park



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **18370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 6, 2018**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1218%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 5, 2020** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
TYLER ROBY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/10/2020

NUMBER 200057

EXPIRES 1/10/2022

MO 580-0771 (6-10)


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ROBY, TYLER
Permit No 200057
Date Issued 1/10/2020 **Date Expires** 1/10/2022

