

RECEIVED

By Tracy Crews at 9:24 am, Jan 24, 2020

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097460 PRINTER SN 096.3580.957 DATE OF INSPECTION 01/22/2020

LOCATION OF INSTRUMENT (STREET AND CITY) 5434 S. Tower Drive, Battlefield, MO 65619 TIME OF INSPECTION 5:27 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER Guth Laboratories LOT # 19160 EXP. DATE 07/09/2021 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2251 SIMULATOR EXP DATE 09/18/2020 CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098

TEST 2 .099

TEST 3 .101

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 | (0-.04) 1 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | (OVER .19) 1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICERSIGNATURE PRINT NAME Ronald M Payne
TYPE II PERMIT NUMBER/EXPIRATION DATE 290164/08-01-2021 TELEPHONE NUMBER (417) 890-9876**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00847

Temp Date Time ^{s/} 210L

VOID: RFI
12 01/22/20 17:30

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

R. Payne / 290164

Location

Buttfield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00846

Temp Date Time ^{s/} 210L

Air Blank:
01/22/20 17:27 .000
Subject Test: Auto
23 01/22/20 17:27 .000

Subject Name

Blank Test

Subject I.D.

Operator Name, I.D.

R. Payne / 290164

Location

Buttfield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00848

Temp Date Time ^{s/} 210L

Air Blank:
01/22/20 17:32 .000
Calibration Check:
24 01/22/20 17:32 .098

Subject Name

ALL check

Subject I.D.

Operator Name, I.D.

R. Payne / 290164

Location

Buttfield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00849

Temp Date Time ^{s/} 210L

Air Blank:
01/22/20 18:01 .000
Calibration Check:
25 01/22/20 18:01 .099

Subject Name

ALL check

Subject I.D.

Operator Name, I.D.

R. Payne / 290164

Location

Buttfield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00850

Temp Date Time ^{s/} 210L

Air Blank:
01/22/20 18:03 .000
Calibration Check:
26 01/22/20 18:03 .101

Subject Name

ALL check

Subject I.D.

Operator Name, I.D.

R. Payne / 290164

Location

Buttfield P.D.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19160** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN04271602** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

RONALD M PAYNE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/1/2019

NUMBER 290164

EXPIRES 8/1/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PAYNE, RONALD
 Permit No 290164
 Date Issued 8/1/2019 Date Expires 8/1/2021

