

RECEIVED

By Tracy Crews at 10:03 am, Mar 30, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097460 PRINTER SN 096.3580.957 DATE OF INSPECTION 03/29/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 5434 S. Tower Drive, Battlefield, MO 65619 TIME OF INSPECTION 4:51 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

[X] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

[X] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

[X] PRINTER WORKING PROPERLY

[X] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

[X] SIMULATOR SOLUTION [] COMPRESSED ETHANOL-GAS MIXTURE

[X] STANDARD SUPPLIER Guth Laboratories LOT # 19160 EXP. DATE 07/09/2021

[X] SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2251 SIMULATOR EXP DATE 09/18/2020

[X] CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [X] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
[] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
[] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101

TEST 2 .103

TEST 3 .101

[X] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE [Signature] PRINT NAME Ronald M Payne
TYPE II PERMIT NUMBER/EXPIRATION DATE 290164/08-01-2021 TELEPHONE NUMBER (417) 890-9876

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00868

Temp Date Time ^{s/} 218L

Air Blank:
03/29/20 16:51 .000
Subject Test: Auto
19 03/29/20 16:51 .000

Subject Name
Blank Test
Subject I.D.

Operator Name, I.D.
R. Payne / 290164
Location
Battlefield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD -00870

Temp Date Time ^{s/} 218L

Air Blank:
03/29/20 16:56 .000
Calibration Check:
20 03/29/20 16:56 .181

Subject Name
ACC check
Subject I.D.

Operator Name, I.D.
R. Payne / 290164
Location
Battlefield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00872

Temp Date Time ^{s/} 218L

Air Blank:
03/29/20 17:00 .000
Calibration Check:
22 03/29/20 17:00 .181

Subject Name
ACC check
Subject I.D.

Operator Name, I.D.
R. Payne / 290164
Location
Battlefield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00869

Temp Date Time ^{s/} 218L

VOID: RFI
12 03/29/20 16:52

Subject Name
RFI + CST
Subject I.D.

Operator Name, I.D.
R. Payne / 290164
Location
Battlefield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00871

Temp Date Time ^{s/} 218L

Air Blank:
03/29/20 16:58 .000
Calibration Check:
21 03/29/20 16:58 .183

Subject Name
ACC check
Subject I.D.

Operator Name, I.D.
R. Payne / 290164
Location
Battlefield P.D.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19160** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
RONALD M PAYNE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repair and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/1/2019

NUMBER 290164

EXPIRES 8/1/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (PG-

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PAYNE, RONALD
Permit No 290164
Date Issued 8/1/2019 Date Expires 8/1/2021