



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Tracy Crews at 2:53 pm, Feb 11, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 099361	PRINTER SN 097.3584.336	DATE OF INSPECTION 02/11/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 106 S. Folger St, Carrollton	TIME OF INSPECTION 8:43 am
--------------------------------------------------------------------------	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 19370 EXP. DATE 12/09/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIMULATOR SN MP3879 SIMULATOR EXP DATE 08/05/2020

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .099	TEST 2 • .099	TEST 3 • .099
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	<u>0</u>	<u>(0-.04)</u>	<u>0</u>	<u>(.05-.09)</u>	<u>0</u>	<u>(.10-.14)</u>	<u>0</u>	<u>(.15-.19)</u>	<u>0</u>	<u>(OVER .19)</u>	<u>0</u>
----------	----------	----------------	----------	------------------	----------	------------------	----------	------------------	----------	-------------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME James H. Plaggemeyer II
---------------	---------------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 280236 08/14/2020	TELEPHONE NUMBER (660) 542-2200
------------------------------------------------------------	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1199%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

JAMES H PLAGGEMEYER II

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/14/2018

NUMBER 280236

EXPIRES 8/14/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator PLAGGEMEYER II, JAMES
 Permit No 280236
 Date Issued 8/14/2018 Date Expires 8/14/2020

CK1

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00690

Temp Date Time ^{s/} 210L

Air Blank:
02/11/20 08:43 .000
Calibration Check:
23 02/11/20 08:43 .099

Subject Name

Subject I.D.

PLAGGEMEYER #785

Operator Name, I.D.

Location

CK2

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00691

Temp Date Time ^{s/} 210L

Air Blank:
02/11/20 08:44 .000
Calibration Check:
23 02/11/20 08:44 .099

Subject Name

PLAGGEMEYER #785

Operator Name, I.D.

Location

CK3

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00692

Temp Date Time ^{s/} 210L

Air Blank:
02/11/20 08:47 .000
Calibration Check:
24 02/11/20 08:47 .099

Subject Name

Subject I.D.

PLAGGEMEYER #785

Operator Name, I.D.

Location

RFI

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00693

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/11/20 08:48

Subject Name

Subject I.D.

PLAGGEMEYER #785

Operator Name, I.D.

Location

TEST

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00694

Temp Date Time ^{s/} 210L

Air Blank:
02/11/20 08:49 .000
Subject Test: Auto
25 02/11/20 08:49 .000

Subject Name

Subject I.D.

PLAGGEMEYER #785

Operator Name, I.D.

Location