

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00742

Temp Date Time ^{s/} 210L

Air Blank:
12/05/20 06:13 .000
Calibration Check:
20 12/05/20 06:13 .103

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00743

Temp Date Time ^{s/} 210L

Air Blank:
12/05/20 06:16 .000
Calibration Check:
21 12/05/20 06:16 .103

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00744

Temp Date Time ^{s/} 210L

Air Blank:
12/05/20 06:18 .000
Calibration Check:
22 12/05/20 06:18 .102

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00745

Temp Date Time ^{s/} 210L

VOID: RFI
12 12/05/20 06:19

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 099361
Version no: 532B

TEST RECORD 00746

Temp Date Time ^{s/} 210L

Air Blank:
12/05/20 06:21 .000
Subject Test: Auto
23 12/05/20 06:21 .000

Subject Name

Blank Test

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **20190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **April 8, 2020**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 6, 2022** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

