



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 8:14 am, Sep 14, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100284	NAME OF AGENCY Columbia Police Department	DATE OF INSPECTION 09/12/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 600 East Walnut Columbia		TIME OF INSPECTION 5:18 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG020401 EXP. DATE 07/22/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .098 TEST 2 ← .098 TEST 3 ← .098

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 3 (0-.04) 0 (.05-.09) 0 (.10-.14) 1 (.15-.19) 1 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Monthly Maintenance

INSPECTING OFFICER

SIGNATURE ▶ <i>Mark D. Hoehne</i>	PRINT NAME Mark D. Hoehne
TYPE II PERMIT NUMBER/EXPIRATION DATE 200187 06/15/2022	TELEPHONE NUMBER (573-874-7585)

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00186

Temp Date Time ^{9/} 210L

Air Blank:
09/12/20 05:21 .000
Calibration Check:
24 09/12/20 05:21 .098

Subject Name

Test # 1

Subject I.D.

Hoehne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00187

Temp Date Time ^{9/} 210L

Air Blank:
09/12/20 05:22 .000
Calibration Check:
24 09/12/20 05:22 .098

Subject Name

Test # 2

Subject I.D.

Hoehne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00188

Temp Date Time ^{9/} 210L

Air Blank:
09/12/20 05:24 .000
Calibration Check:
24 09/12/20 05:24 .098

Subject Name

Test # 3

Subject I.D.

Hoehne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00189

Temp Date Time ^{9/} 210L

VOID: RFI
12 09/12/20 05:25

Subject Name

RFI Check

Subject I.D.

Hoehne, Mark D. 2078
Operator Name, I.D.

Location

AS IV Serial no: 100284
Version no: 532B

TEST RECORD 00190

Temp Date Time ^{9/} 210L

Air Blank:
09/12/20 05:26 .000
Subject Test: Auto
25 09/12/20 05:26 .000

Subject Name

Self Test

Subject I.D.

Hoehne, Mark D. 2078
Operator Name, I.D.

Location



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 22-Jul-2020

Lot # AG020401 **Model** 108cacc

Exp. Date

22-Jul-2022

Cyl. Type

108

Component

Ethanol
Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

Concentration

392.1 ppm

EB0010570

259.8 ppm

EB0010285

208.0 ppm

EB0010561

103.6 ppm

EB0010681

52.12 ppm

RGM Serial No.

EB0010603

Concentration

393.0 ppm

EB0010559

258.2 ppm

EB0010595

208.3 ppm

EB0010562

104.2 ppm

EB0010579

52.81 ppm

CRM Serial No.

CC434668

Concentration

800.0 ppm

CC234503

253.0 ppm

CRM Serial No.

0056649

Concentration

390.1 ppm

0056662

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2020.07.22 18:29:45 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07

**PERMIT
TYPE II**

MARK D HOEHNE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/15/2020

NUMBER **200187**

EXPIRES 6/15/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (16-10)

MO SFD 0771 (6-19)

 **STATE OF MISSOURI**
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri

Operator **HOEHNE, MARK**
Permit No **200187**
Date Issued **6/15/2020** Date Expires **6/15/2022**

