



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100284	NAME OF AGENCY Columbia Police Department	DATE OF INSPECTION 11/11/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 600 East Walnut Columbia		TIME OF INSPECTION 7:30 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG020401</u> EXP. DATE <u>07/22/2022</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100	TEST 2 ← .099	TEST 3 ← .099
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS <b>2</b>	(0-.04) <b>0</b>	(.05-.09) <b>0</b>	(.10-.14) <b>0</b>	(.15-.19) <b>2</b>	(OVER .19) <b>1</b>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

# Monthly Maintenance

**INSPECTING OFFICER**

SIGNATURE ▶ <i>Mark D. Hoehne</i> 2078	PRINT NAME Mark D. Hoehne
TYPE II PERMIT NUMBER/EXPIRATION DATE 200187 06/15/2022	TELEPHONE NUMBER (573-874)-7585

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 100284  
Version no: 532B

TEST RECORD 00206

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
11/11/20 19:31 .000  
Calibration Check:  
22 11/11/20 19:31 .100

Subject Name

Test #1

Subject I.D.

Hoehne, Mark D. 2078  
Operator Name, I.D.

Location

AS IV Serial no: 100284  
Version no: 532B

TEST RECORD 00207

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
11/11/20 19:33 .000  
Calibration Check:  
23 11/11/20 19:33 .099

Subject Name

Test #2

Subject I.D.

Hoehne, Mark D. 2078  
Operator Name, I.D.

Location

AS IV Serial no: 100284  
Version no: 532B

TEST RECORD 00208

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
11/11/20 19:35 .000  
Calibration Check:  
23 11/11/20 19:35 .099

Subject Name

Test #3

Subject I.D.

Hoehne, Mark D. 2078  
Operator Name, I.D.

Location

AS IV Serial no: 100284  
Version no: 532B

TEST RECORD 00209

Temp Date Time <sup>g/</sup> 210L

VOID: RFI  
12 11/11/20 19:37

Subject Name

RFI Check

Subject I.D.

Hoehne, Mark D. 2078  
Operator Name, I.D.

Location

AS IV Serial no: 100284  
Version no: 532B

TEST RECORD 00210

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
11/11/20 19:38 .000  
Subject Test: Auto  
24 11/11/20 19:38 .000

Subject Name

Self Test

Subject I.D.

Hoehne, Mark D. 2078  
Operator Name, I.D.

Location



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 22-Jul-2020

**Lot # AG020401 Model 108cacd**

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
22-Jul-2022	108	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2020.07.22 18:29:45 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

**Approved for Release:**   
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**

TYPE II

MARK D HOEHNE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

### ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/15/2020

NUMBER 200187

EXPIRES 6/15/2022

2020 01/10 (0-19)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (15-10)

**STATE OF MISSOURI**  
**DEPARTMENT OF HEALTH AND SENIOR SERVICES**  
**BREATH ALCOHOL PROGRAM**

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator **HOEHNE, MARK**  
 Permit No **200187**  
 Date Issued **6/15/2020**    Date Expires **6/15/2022**

