



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100286	PRINTER SN 093.3563.013	DATE OF INSPECTION 1-1-20
LOCATION OF INSTRUMENT (STREET AND CITY) 602 S. Franklin Cuba		TIME OF INSPECTION 6:48 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS-OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth LOT # 19070 EXP. DATE 2-25-21
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34° SIMULATOR SN MP3872 SIMULATOR EXP DATE 4-18-20

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ¹⁰⁰ .100	TEST 2 ¹⁰⁰ .099	TEST 3 ¹⁰⁰ .099
-----------------------------------	-----------------------------------	-----------------------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	(OVER .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>Michael Centunzi</i>	PRINT NAME Michael Centunzi
TYPE II PERMIT NUMBER/EXPIRATION DATE 280179/5-7-2020	TELEPHONE NUMBER 573-885-7979

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

98 10 2019 001 280179
CANTINI / 280179

1181 01/201 20121

98 10 2019 001 280179
CANTINI / 280179

Blank

Test 1

Cantini / 280179

602 S. Franklin

Cuba

98 10 2019 001 280179
CANTINI / 280179

1181 01/201 20121

98 10 2019 001 280179
CANTINI / 280179

Blank

Test 2

Cantini / 280179

602 S. Franklin

Cuba

98 10 2019 001 280179
CANTINI / 280179

1181 01/201 20121

98 10 2019 001 280179
CANTINI / 280179

Blank

Test 3

Cantini / 280179

602 S. Franklin

Cuba

98 10 2019 001 280179
CANTINI / 280179

1181 01/201 20121

98 10 2019 001 280179
CANTINI / 280179

Blank

RFI

Cantini / 280179

602 S. Franklin

Cuba



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

MICHAEL S CENTUNZI

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/7/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 280179

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 5/7/2020

LAB-4 (06-10)

590-0771 (6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator: CENTUNZI, MICHAEL
 Permit No: 280179
 Date Issued 5/7/2018 Date Expires 5/7/2020



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19070** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **February 26, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1212%** (w/vol) ethyl alcohol. The expiration date for this lot number is **February 25, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466
 Randall W. Williams, MD, FACOG
 Director



Michael L. Parson
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP3872 Manufacturer: Guth
 Model Number: 12V500
 Agency: CUBA PD
 Agency Address: 602 S FRANKLIN, CUBA, MO 65453

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00890 Bias: 0.00
 Uncertainty: 0.02
 Date of Certification: 1/31/2019 Date of Expiration: 1/31/2020

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.01	.03

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 4/18/2019
 Certification Expiration: 4/18/2020
 Simulator testing technician: S. GARY

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
 Certification No: MP3872_4182019

X

DHSS BAP Scientist Approving