



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 100289	PRINTER SN 093.3563.017	DATE OF INSPECTION 06/19/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 9701 Marion Park Dr, Kansas City	TIME OF INSPECTION 6:49 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u>	LOT # <u>AG905605</u> EXP. DATE <u>02/25/2021</u>

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .080	TEST 2 • .080	TEST 3 • .079
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	3	(0-.04)	1	(.05-.09)	1	(.10-.14)	2	(.15-.19)	1	(OVER .19)	3
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER

SIGNATURE <i>P.O. Jordan Infranca 5670</i>	PRINT NAME Jordan Infranca
---	-------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 290130/ 06/21/2021	TELEPHONE NUMBER (816) 382-5897
---	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 100289
Version no: 532B

TEST RECORD 00369

Temp Date Time 210L

Air Blank:
06/19/20 18:43 .000
Calibration Check:
22 06/19/20 18:43 .000

Subject Name

Test # 1

Subject I.D.

Infranca 290130

Operator Name, I.D.

KCPD Traffic unit

Location

9701 Marion Park

Drive, KCMO

AS IV Serial no: 100289
Version no: 532B

TEST RECORD 00370

Temp Date Time 210L

Air Blank:
06/19/20 18:45 .000
Calibration Check:
23 06/19/20 18:45 .000

Subject Name

Test # 2

Subject I.D.

Infranca 290130

Operator Name, I.D.

KCPD Traffic unit

Location

9701 Marion Park

Drive Kemo

AS IV Serial no: 100289
Version no: 532B

TEST RECORD 00371

Temp Date Time 210L

Air Blank:
06/19/20 18:48 .000
Calibration Check:
24 06/19/20 18:48 .079

Subject Name

Test # 3

Subject I.D.

Infranca 290130

Operator Name, I.D.

9701 Marion Park

Location

Drive KCMO

KCPD Traffic unit

AS IV Serial no: 100289
Version no: 532B

TEST RECORD 00372

Temp Date Time 210L

VOID: RFI
12 06/19/20 18:49

Subject Name

RFI

Subject I.D.

Infranca 290130

Operator Name, I.D.

9701 Marion Park

Location

Drive, KCMO

KCPD Traffic unit



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcohol content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 806.111 through 806.119 RSMo.

DATE 6/21/2019

NUMBER 290130

EXPIRES 6/21/2021

ISO 9100:2015 (I-19)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAW-4 (9-19)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named operator is authorized to operate an instrument that is used to determine the alcohol content of the alcohol content in breath from a sample of expired air.

Operator: INFRANCA, JORDAN
Permit No: 290130
Date Issued: 6/21/2019 Date Expires: 6/21/2021

Airgas

Airgas USA LLC (LAB)
3600 Bernard Street
St. Louis, Mo, 63103
Ph: (314) 633-3100
Fax: (314) 633-7320

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 28-Feb-2019

Lot # AG905605 Model 1080acd

Exp. Date	Cvl. Type	Component	Certified Concentration
26-Feb-2021	108	Ethanol Nitrogen	0.082 ± 0.002 BrAC (223 ppm) Balance

Certification Traceable to N.I.S.T, RGM and to GRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010561	382.1 ppm	EB0010563	393.0 ppm
EB0010570	259.0 ppm	EB0010559	258.2 ppm
EB0010286	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

GRM Serial No.	Concentration	GRM Serial No.	Concentration
CG434663	800.0 ppm	0056640	390.1 ppm
QC234603	283.0 ppm	0056602	160.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2019.02.28 11:27:05 -0500
Reason: Dry gas alcohol and calibration of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:
Rod Marsala

ISO 17025:2005 A2LA accredited, Certificate Number 3002.06
ISO 17034:2016 A2LA accredited, Certificate Number 3002.07