



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED REPORT #7
 By Tracy Crews at 7:53 am, Sep 25, 2020

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102454	PRINTER SN 095.3583.044	DATE OF INSPECTION 09/15/2020
LOCATION OF INSTRUMENT (STREET AND CITY) Monett Police Department, 1901 E. Cleveland Ave, Monett, MO, 65708		TIME OF INSPECTION

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>Ag015503</u> EXP. DATE <u>06/22/2022</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .099	TEST 2 ← .099	TEST 3 ← .099
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14) 3	(.15-.19)	(OVER .19)
----------	---------	-----------	----------------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Trent Gold
TYPE, IL PERMIT NUMBER/EXPIRATION DATE 290051 3/1/2021	TELEPHONE NUMBER 417-235-4241

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00359

Temp Date Time ^{a/} 210L

Air Blank:
09/15/20 15:38 .000
Calibration Check:
18 09/15/20 15:38 .099

Subject Name

Test 1
Subject I.D.

Operator Name, I.D.

Trent Gold 290051
Location
Monett PD

1901 E. Cleveland Ave
Monett mo 65708

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00360

Temp Date Time ^{a/} 210L

Air Blank:
09/15/20 15:39 .000
Calibration Check:
19 09/15/20 15:39 .099

Subject Name

Test 2
Subject I.D.

Operator Name, I.D.

Trent Gold 290051
Location
Monett PD

1901 E. Cleveland Ave
Monett mo 65708

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00361

Temp Date Time ^{a/} 210L

Air Blank:
09/15/20 15:41 .000
Calibration Check:
20 09/15/20 15:41 .099

Subject Name

Test 3
Subject I.D.

Operator Name, I.D.

Trent Gold 290051
Location
Monett PD.

1901 E. Cleveland Ave
Monett mo 65708

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00362

Temp Date Time ^{a/} 210L

VOID: RFI
12 09/15/20 15:42

Subject Name

RFI
Subject I.D.

Operator Name, I.D.

Trent Gold 290051
Location
Monett PD.

1901 E. Cleveland Ave
Monett mo 65708

AS IV Serial no: 102454
Version no: 532B

TEST RECORD 00363

Temp Date Time ^{a/} 210L

Air Blank:
09/15/20 15:43 .000
Subject test: Auto
22 09/15/20 15:43 .000

Subject Name

Operational test
Subject I.D.

Operator Name, I.D.

Trent Gold 290051
Location
Monett PD.

1901 E. Cleveland Ave
Monett mo 65708



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 8-Jun-2020

Lot # AG015503 Model 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
3-Jun-2022	108	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2020.06.10 14:11:39 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
TRENT GOLD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/1/2019

NUMBER 290051

EXPIRES 3/1/2021

MO 580-0771 (5-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (RS-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GOLD, TRENT
Permit No 290051
Date Issued 3/1/2019 **Date Expires** 3/1/2021