

RECEIVED

001/004

By Tracy Crews at 9:52 am, Apr 13, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102458	PRINTER SN 093.3578.252	DATE OF INSPECTION 04/11/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 116 W. Brooks St. Brookfield, MO 64628		TIME OF INSPECTION 09:22 p.m.

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo LOT # 18001 EXP. DATE 07/31/2020

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2763 SIMULATOR EXP DATE 01/08/2021

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 = 0.099

TEST 2 = 0.098

TEST 3 = 0.098

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

The instrument meets Department of Health standards.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Coy Hunt
TYPE II PERM. NUMBER/EXPIRATION DATE 280195-05/24/2020	TELEPHONE NUMBER (660) 258-3385

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 102452
Version no: 532B

TEST RECORD 00652

Temp Date Time 21:26

Air Blank: 04/11/20 21:26 .000
Calibration Check: 24 04/11/20 21:26 .000

Subject Name

RFI

Subject I.D.

Operator Name: I.D.

28019 8

Location

RPD

AS IV Serial no: 102459
Version no: 532B

TEST RECORD 00651

Temp Date Time 21:26

Air Blank: 04/11/20 21:26 .000
Calibration Check: 24 04/11/20 21:26 .000

Subject Name

Simulator #3

Subject I.D.

Operator Name: I.D.

28019 8

Location

RPD

AS IV Serial no: 102458
Version no: 532B

TEST RECORD 00650

Temp Date Time 21:24

Air Blank: 04/11/20 21:24 .000
Calibration Check: 23 04/11/20 21:24 .000

Subject Name

Simulator #2

Subject I.D.

Operator Name: I.D.

28019 8

Location

RPD

AS IV Serial no: 102453
Version no: 532B

TEST RECORD 00659

Temp Date Time 21:22

Air Blank: 04/11/20 21:22 .000
Calibration Check: 23 04/11/20 21:22 .000

Subject Name

Simulator #1

Subject I.D.

Operator Name: I.D.

28019 8

Location

RPD

RepCo

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
888-828-0227

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 18001
EXPIRATION DATE: July 31, 2020 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 18001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1206 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 1, 2018
The expiration date for this lot number is July 31, 2020 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
COY HUNT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/24/2018

NUMBER 280198

EXPIRES 5/24/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4-(R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator HUNT, COY
Permit No 280198
Date Issued 5/24/2018 Date Expires 5/24/2020