

Subject Name

AS IV Serial no: 102463
Version no: 532B

TEST RECORD 00272

Temp Date Time 210L ^{s/}

Air Blank:
05/24/20 09:37 .000
Calibration Check:
24 05/24/20 09:37 .101

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

E.J. GANZME 200019

Location

MADISON CO. S.O.

AS IV Serial no: 102463
Version no: 532B

TEST RECORD 00273

Temp Date Time 210L ^{s/}

Air Blank:
05/24/20 09:39 .000
Calibration Check:
24 05/24/20 09:39 .100

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

E.J. GANZME 200019

Location

MADISON CO. S.O.

AS IV Serial no: 102463
Version no: 532B

TEST RECORD 00274

Temp Date Time 210L ^{s/}

Air Blank:
05/24/20 09:40 .000
Calibration Check:
25 05/24/20 09:40 .099

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

E.J. GANZME 200019

Location

MADISON CO. S.O.

AS IV Serial no: 102463
Version no: 532B

TEST RECORD 00275

Temp Date Time 210L ^{s/}

VOID: RFI
12 05/24/20 09:41

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

E.J. GANZME

Location

MADISON CO. S.O.

CERTIFICATE OF ANALYSIS**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.****LOT NUMBER: 19002****EXPIRATION DATE: October 16, 2021 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
ERIC J GANIME

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/6/2020

NUMBER 200019

EXPIRES 1/6/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GANIME, ERIC
 Permit No 200019
 Date Issued 1/6/2020 Date Expires 1/6/2022

