



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 104642	NAME OF AGENCY O'Fallon PD	DATE OF INSPECTION 9-25-2020
LOCATION OF INSTRUMENT (STREET AND CITY) 1019 Bryan Road, O'Fallon, MO		TIME OF INSPECTION 1120

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u> LOT # <u>AG919902</u> EXP. DATE <u>07-18-2021</u>	
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____	

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input checked="" type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .079	TEST 2 ← .079	TEST 3 ← .078
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	(OVER .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Time adjusted (-3 minutes)

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Chris Sinnokrak
TYPE II PERMIT NUMBER/EXPIRATION DATE 290008 / 1-11-2021	TELEPHONE NUMBER (636)240-3200

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 104642
Version no: 532B

TEST RECORD 00762

Temp Date Time ^{s/} 210L

Air Blank:
09/25/20 10:40 .000
Calibration Check:
21 09/25/20 10:40 .079

Subject Name
TEST 1
Subject I.D.

Operator Name, I.D.
SINNOGRAK 334
Location

AS IV Serial no: 104642
Version no: 532B

TEST RECORD 00763

Temp Date Time ^{s/} 210L

Air Blank:
09/25/20 10:42 .000
Calibration Check:
22 09/25/20 10:42 .079

Subject Name
TEST 2
Subject I.D.

Operator Name, I.D.
SINNOGRAK 334
Location

AS IV Serial no: 104642
Version no: 532B

TEST RECORD 00764

Temp Date Time ^{s/} 210L

Air Blank:
09/25/20 10:44 .000
Calibration Check:
23 09/25/20 10:44 .078

Subject Name
TEST ?
Subject I.D.

Operator Name, I.D.
SINNOGRAK 334
Location

AS IV Serial no: 104642
Version no: 532B

TEST RECORD 00765

Temp Date Time ^{s/} 210L

VOID: RFI
12 09/25/20 10:46

Subject Name
RFI
Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 104642
Version no: 532B

TEST RECORD 00766

Temp Date Time ^{s/} 210L

Air Blank:
09/25/20 10:47 .000
Subject Test: Auto
23 09/25/20 10:47 .000

Subject Name
SOBER TEST
Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 22-Jul-2019

Lot # AG919902 Model 108caccd

<u>Exp. Date</u> 18-Jul-2021	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (218 ppm) Balance
--	--------------------------------	--	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u> EB0010581 EB0010570 EB0010285 EB0010561 EB0010681	<u>Concentration</u> 392.1 ppm 259.8 ppm 208.0 ppm 103.6 ppm 52.12 ppm	<u>RGM Serial No.</u> EB0010603 EB0010559 EB0010595 EB0010562 EB0010579	<u>Concentration</u> 393.0 ppm 258.2 ppm 208.3 ppm 104.2 ppm 52.81 ppm
<u>CRM Serial No.</u> CC434668 CC234503	<u>Concentration</u> 800.0 ppm 253.0 ppm	<u>CRM Serial No.</u> 0056649 0056662	<u>Concentration</u> 390.1 ppm 150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2019.07.22 13:17:50 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
CHRIS SINNOKRAK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/11/2019

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 290008

EXPIRES 1/11/2021

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SINNOKRAK, CHRIS
Permit No 290008
Date Issued 1/11/2019 Date Expires 1/11/2021