



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED  
 07/19/2020 09:41 AM

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105443	PRINTER SN 099.3586.172	DATE OF INSPECTION 05/07/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 100 Municipal Circle Raymore, Missouri 64083	TIME OF INSPECTION 5:17 am
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories, Inc. LOT # 19341 EXP. DATE 11/18/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2256 SIMULATOR EXP DATE 07/26/2020

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .096

TEST 2 ← .098

TEST 3 ← .098

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
 Joshua Giacone

TYPE II PERMIT NUMBER/EXPIRATION DATE  
 200151 04/06/2022

TELEPHONE NUMBER  
 (816) 331-0530

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 105443  
Version no: 532B

TEST RECORD 00940

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
05/07/20 05:18 .000  
Calibration Check:  
21 05/07/20 05:18 .096

Subject Name

Maintenance

Subject I.D.

J. Giaccone #919

Operator Name, I.D.

Patmore P.D.

Location

AS IV Serial no: 105443  
Version no: 532B

TEST RECORD 00941

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
05/07/20 05:21 .000  
Calibration Check:  
21 05/07/20 05:21 .098

Subject Name

Maintenance

Subject I.D.

J. Giaccone #919

Operator Name, I.D.

Patmore P.D.

Location

AS IV Serial no: 105443  
Version no: 532B

TEST RECORD 00942

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
05/07/20 05:23 .000  
Calibration Check:  
22 05/07/20 05:23 .098

Subject Name

Maintenance

Subject I.D.

J. Giaccone #919

Operator Name, I.D.

Patmore P.D.

Location

AS IV Serial no: 105443  
Version no: 532B

TEST RECORD 00943

Temp Date Time <sup>g/</sup> 210L

VOID: RFI  
12 05/07/20 05:25

Subject Name

Maintenance

Subject I.D.

J. Giaccone #919

Operator Name, I.D.

Patmore P.D.

Location



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19341** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **November 20, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1211%** (w/vol) ethyl alcohol. The expiration date for this lot number is **November 18, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** SD2256      **Manufacturer:** Guth  
**Model Number:** 10-4D  
**Agency:** RAYMORE PD  
**Agency Address:** 100 MUNICIPAL CIRCLE, RAYMORE, MO 64083

## NIST THERMOMETER INFORMATION

**Serial Number:** 17KMM00689      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 12/11/2018      **Date of Expiration:** 12/11/2019

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	33.99	.03

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 7/26/2019  
**Certification Expiration:** 7/26/2020  
**Simulator testing technician:** D. DEBOARD

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** B. LUTMER  
**Certification No:** SD2256\_7262019

X

DHSS BAP Scientist Approving



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JOSHUA B GIACONE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2020

NUMBER 200151

EXPIRES 4/6/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator GIACONE, JOSHUA  
 Permit No 200151  
 Date Issued 4/6/2020 Date Expires 4/6/2022

