

RECEIVED

By Tracy Crews at 11:23 am, Jul 07, 2020



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105446 PRINTER SN 099.3586.164 DATE OF INSPECTION 05/31/2020

LOCATION OF INSTRUMENT (STREET AND CITY) 401 N. Bynum, Lone Jack, MO. 64070 TIME OF INSPECTION 01:59 PM

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 19180 EXP. DATE 07/09/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2238 SIMULATOR EXP DATE 05/28/2021

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.100%

TEST 2 ← 0.098%

TEST 3 ← 0.098%

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 1 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary):

INSPECTING OFFICER

SIGNATURE *[Signature]* #372

PRINT NAME Zach Argetsinger

TYPE II PERMIT NUMBER/EXPIRATION DATE 290072 03/18/2021

TELEPHONE NUMBER (816) 697-2417

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00780

Temp Date Time 210L
Air Blank: 05/31/20 14:14 .000

Calibration Check: 27 05/31/20 14:14 .998

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Argetsinger 372

Location

Y01N. Bynum

Lone Jack, MO

TEST RECORD 00779

Temp Date Time 210L

Air Blank: 05/31/20 14:12 .000

Calibration Check: 26 05/31/20 14:12 .998

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Argetsinger 372

Location

Y01N. Bynum

Lone Jack, MO

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00778

Temp Date Time 210L

Air Blank: 05/31/20 14:01 .000

Calibration Check: 25 05/31/20 14:01 .100

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Argetsinger 372

Location

Y01N. Bynum

Lone Jack, MO

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00777

Temp Date Time 210L

Air Blank: 05/31/20 13:59 .000

Calibration Check: 25 05/31/20 13:59 .000

Subject Name

Diag Readout Test

Subject I.D.

Operator Name, I.D.

Argetsinger 372

Location

Y01N. Bynum

Lone Jack, MO

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00781

Temp Date Time 210L

VOID: RFI
12 05/31/20 14:16

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Argetsinger 372

Location

Y01N. Bynum

Lone Jack, MO



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 19160 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on July 10, 2019, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1202% (w/vol) ethyl alcohol. The expiration date for this lot number is July 9, 2021 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

ZACHARY A ARGETSINGER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/18/2019

NUMBER 290072

EXPIRES 3/18/2021

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAR-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ARGETSINGER, ZACHARY
Permit No 290072
Date Issued 3/18/2019 Date Expires 3/18/2021