

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00798
Temp Date Time 210L

Air Blank: 09/20/20 14:14 .000
Calibration Check: 27 09/20/20 14:14 .099

Subject Name
Test # 3
Subject I.D.

Operator Name, I.D.
Argetsinger
Location
Y01N.Bynum

LoneJack,MO

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00797
Temp Date Time 210L

Air Blank: 09/20/20 14:12 .000
Calibration Check: 27 09/20/20 14:12 .099

Subject Name
Test # 2
Subject I.D.

Operator Name, I.D.
Argetsinger
Location
Y01N.Bynum

LoneJack,MO

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00796
Temp Date Time 210L

Air Blank: 09/20/20 14:10 .000
Calibration Check: 26 09/20/20 14:10 .099

Subject Name
Test # 1
Subject I.D.

Operator Name, I.D.
Argetsinger
Location
Y01N.Bynum

LoneJack,MO

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00795
Temp Date Time 210L

Air Blank: 09/20/20 14:08 .000
Calibration Check: 26 09/20/20 14:08 .080

Subject Name
Dig Readout
Subject I.D.

Operator Name, I.D.
Argetsinger
Location
Y01N.Bynum

LoneJack,MO

AS IV Serial no: 105446
Version no: 532B

TEST RECORD 00799

Temp Date Time 210L

VOID: RFI
12 09/20/20 14:16

Subject Name
RFI Test
Subject I.D.

Operator Name, I.D.

Argetsinger
Location
Y01N.Bynum

LoneJack,MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19160** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **July 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **July 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number **FN04271602** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

ZACHARY A ARGETSINGER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/18/2019

NUMBER 290072

EXPIRES 3/18/2021

MO 580-0771 (6-10)

[Signature]
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (05-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ARGETSINGER, ZACHARY
Permit No 290072
Date Issued 3/18/2019 Date Expires 3/18/2021

