



RECEIVED

By Stephen Wilson at 9:10 am, May 26, 2020

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107971	PRINTER SN 099.3586.826	DATE OF INSPECTION 05/24/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 17010 Highway 87, Boonville		TIME OF INSPECTION 6:16 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Repco LOT # 19002 EXP. DATE 10/16/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN MP2202 SIMULATOR EXP DATE 01/31/2021

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101

TEST 2 .098

TEST 3 .100

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Tested within DHSS standards.

INSPECTING OFFICER

SIGNATURE
Trooper A.M. Richerson 759

PRINT NAME
Trooper A.M. Richerson

TYPE & PERMIT NUMBER/EXPIRATION DATE
290186 - 8/19/2021

TELEPHONE NUMBER
(573) 751-1000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 107971
Version no: 532B

TEST RECORD 00563

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:
05/24/20 18:25 .000
Calibration Check:
23 05/24/20 18:25 .101

Subject Name

Test 1

Subject I.D.

1234

Operator Name, I.D.

Richerson 290186

Location

17010 Hwy 87
Boonville

AS IV Serial no: 107971
Version no: 532B

TEST RECORD 00564

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:
05/24/20 18:27 .000
Calibration Check:
24 05/24/20 18:27 .090

Subject Name

Test 2

Subject I.D.

1234

Operator Name, I.D.

Richerson 290186

Location

17010 Hwy 87
Boonville

AS IV Serial no: 107971
Version no: 532B

TEST RECORD 00565

Temp	Date	Time	a/ 210L
------	------	------	------------

Air Blank:
05/24/20 18:29 .000
Calibration Check:
25 05/24/20 18:29 .100

Subject Name

Test 3

Subject I.D.

1234

Operator Name, I.D.

Richerson 290186

Location

17010 Hwy 87
Boonville

AS IV Serial no: 107971
Version no: 532B

TEST RECORD 00566

Temp	Date	Time	a/ 210L
------	------	------	------------

VOID: RFI
12 05/24/20 18:30

Subject Name

RFI Test

Subject I.D.

1234

Operator Name, I.D.

Richerson 290186

Location

17010 Hwy 87
Boonville

CERTIFICATE OF ANALYSIS**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.****LOT NUMBER: 19002****EXPIRATION DATE: October 16, 2021 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

ANTHONY M RICHERSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/19/2019

NUMBER 290186

EXPIRES 8/19/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **RICHERSON, ANTHONY**
Permit No **290186**
Date Issued **8/19/2019** Date Expires **8/19/2021**

