



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107973	PRINTER SN 099.3586.816	DATE OF INSPECTION 09/10/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 690 Chesterfield Pkwy W, Chesterfield MO		TIME OF INSPECTION 1:13 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER ILMO LOT # 1781908011 EXP. DATE 08/05/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ 0.080

TEST 2 ➡ 0.079

TEST 3 ➡ 0.078

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	5	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within the Missouri Department of Health Specifications.

INSPECTING OFFICER

SIGNATURE

PRINT NAME
G. Rupp #264

TYPE II PERMIT NUMBER/EXPIRATION DATE
290094 / 04/22/2021

TELEPHONE NUMBER
636-537-3000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 107973
Version no: 532B

TEST RECORD 00751

Temp Date Time 210L

Air Blank:
09/10/20 13:16 .000
Calibration Check:
24 09/10/20 13:16 .078

Subject Name

Test # 3

Subject I.D.

G. RUPP # 264

Operator Name, I.D.

290094

Location

690 Chestnutfield

Pkwy W

AS IV Serial no: 107973
Version no: 532B

TEST RECORD 00750

Temp Date Time 210L

Air Blank:
09/10/20 13:15 .000
Calibration Check:
23 09/10/20 13:15 .079

Subject Name

Test # 2

Subject I.D.

G. RUPP # 264

Operator Name, I.D.

290094

Location

690 Chestnutfield

Pkwy W

AS IV Serial no: 107973
Version no: 532B

TEST RECORD 00749

Temp Date Time 210L

Air Blank:
09/10/20 13:13 .000
Calibration Check:
22 09/10/20 13:13 .080

Subject Name

Test # 1

Subject I.D.

G. RUPP # 264

Operator Name, I.D.

290094

Location

690 Chestnutfield

Pkwy W

AS IV Serial no: 107973
Version no: 532B
TEST RECORD 00752
Temp Date Time 210L
VOID: REI
12 09/10/20 13:17
Subject Name
RFI Test
Subject I.D.
G. RUPP # 264
Operator Name, I.D.
290094
Location
690 Chestnutfield
Pkwy W



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 12199
Part #: BAC108L080I
Cylinder Size: 108L
Lot Number: 17819080I1
Expiration: 8/5/2021

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 108 Liters @ 1200 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	218 ppm	+/- 0.002 BAC (G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		


***Traceable to:**

Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).


Specialty Gas Lab Tech

07-05-19
Date



The calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2005 Accredited Laboratory



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
GREGORY RUPP

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2019

NUMBER 290094

EXPIRES 4/22/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator RUPP, GREGORY
Permit No 290094
Date Issued 4/22/2019 **Date Expires** 4/22/2021

