


 MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107977	PRINTER SN 096.3580.877	DATE OF INSPECTION 06/02/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 301 S MAIN ST, EXCELSIOR SPRINGS, MO 64024		TIME OF INSPECTION 12:16 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

 DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

 TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

 PRINTER WORKING PROPERLY

 TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**
 SIMULATOR SOLUTION

 COMPRESSED ETHANOL-GAS MIXTURE

 STANDARD SUPPLIER GUTH LOT # 19730 EXP. DATE 12/10/2021
 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN MP3564 SIMULATOR EXP DATE 06/18/2020
 CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .100

TEST 2 → .099

TEST 3 → .098

 RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	14	(.15-.19)	0	(OVER .19)	1
----------	---	---------	---	-----------	---	-----------	----	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME

SGT BRIAN K KENNEDY

TYPE II PERMIT NUMBER/EXPIRATION DATE

290040 02/20/2021

TELEPHONE NUMBER

(816) 630-2000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS IV Serial no: 107977  
Version no: 502B

TEST RECORD 00092  
Temp Date Time 210L

UOIR: RTI  
12 06/02/20 12:29  
Subject Name

SIMULATOR  
Subject I.D.

Operator Name, I.D.

KENNEDY 911  
Location

301 S MAIN ST

EX SPG, MO

AS IV Serial no: 107977  
Version no: 502B

TEST RECORD 00093  
Temp Date Time 210L

Air Blank:  
06/02/20 12:32 .000  
Calibration Check:  
25 06/02/20 12:32 .100

Subject Name  
SIMULATOR  
Subject I.D.

Operator Name, I.D.

KENNEDY 911  
Location

301 S MAIN ST

EX SPG, MO

AS IV Serial no: 107977  
Version no: 502B

TEST RECORD 00094  
Temp Date Time 210L

Air Blank:  
06/02/20 12:34 .000  
Calibration Check:  
25 06/02/20 12:34 .099

Subject Name  
SIMULATOR  
Subject I.D.

Operator Name, I.D.

KENNEDY 911  
Location

301 S MAIN ST

EX SPG, MO

AS IV Serial no: 107977  
Version no: 502B

TEST RECORD 00095  
Temp Date Time 210L

Air Blank:  
06/02/20 12:36 .000  
Calibration Check:  
26 06/02/20 12:36 .098

Subject Name  
SIMULATOR  
Subject I.D.

Operator Name, I.D.

KENNEDY 911  
Location

301 S MAIN ST

EX SPG, MO

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **19370** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **December 10, 2019**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1199%** (w/vol) ethyl alcohol. The expiration date for this lot number is **December 9, 2021** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN04271602 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



Missouri Department of Health and Senior Services  
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466  
 Randall W. Williams, MD, FACOG  
 Director



Michael L. Parson  
 Governor

## SIMULATOR CERTIFICATION REPORT

### SIMULATOR INFORMATION

Simulator Serial Number: MP3564      Manufacturer: Guth  
 Model Number: 12V500  
 Agency: EXCELSIOR SPRINGS  
 Agency Address: 301 S MAIN, EXCELSIOR SPRINGS, MO 64024

### NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690      Bias: 0.00  
 Uncertainty: 0.02  
 Date of Certification: 1/31/2019      Date of Expiration: 1/31/2020

### ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

### VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
33.99	34.00	.03

The combined uncertainty is calculated with a k=2 value.

### ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 6/17/2019  
 Certification Expiration: 6/17/2020  
 Simulator testing technician: S. GARY

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER  
 Certification No: MP3564\_6172019

X

DHSS BAP Scientist Approving



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**BRIAN K KENNEDY**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/20/2019

NUMBER 290040

EXPIRES 2/20/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** KENNEDY, BRIAN  
**Permit No** 290040  
**Date Issued** 2/20/2019 **Date Expires** 2/20/2021

