



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By crewst at 10:47 am, Feb 24, 2020

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>107985</u>	PRINTER SN <u>099.3586.820</u>	DATE OF INSPECTION <u>2-25-20</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>200 North Ave, Sparta MO, 65753</u>		TIME OF INSPECTION <u>0920</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>GUTH LAB</u>	LOT # <u>18370</u> EXP. DATE <u>12/5/20</u>
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34°C</u>	SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <u>.101</u>	TEST 2 • <u>.101</u>	TEST 3 • <u>.101</u>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04) <u>1</u>	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	------------------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Changed Time, Got 1 minute behind

INSPECTING OFFICER

SIGNATURE 	PRINT NAME <u>LOREN WYSTROM</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>290056 / 3/1/2021</u>	TELEPHONE NUMBER <u>417242-5511</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II

LOREN NYSTROM

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/1/2019

NUMBER 290056

EXPIRES 3/1/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 560-0771 (6-10)

LAB-4 (R6-10)




STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator NYSTROM, LOREN
 Permit No 290056
 Date Issued 3/1/2019 Date Expires 3/1/2021



AS IV Serial no: 107985
Version no: 532B

TEST RECORD 00890

Temp Date Time ^{s/} 210L

Air Blank:
02/21/20 09:31 .000
Calibration Check:
23 02/21/20 09:31 .101

Subject Name

Test

Subject I.D.

#1

Operator Name, I.D.

L. Nyström ~~615~~

Location

SPD Booking

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 00891

Temp Date Time ^{s/} 210L

Air Blank:
02/21/20 09:33 .000
Calibration Check:
23 02/21/20 09:33 .101

Subject Name

Test

Subject I.D.

#2

Operator Name, I.D.

L. Nyström 615

Location

SPD Booking

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 00892

Temp Date Time ^{s/} 210L

Air Blank:
02/21/20 09:35 .000
Calibration Check:
24 02/21/20 09:35 .101

Subject Name

Test

Subject I.D.

#3

Operator Name, I.D.

L. Nyström 615

Location

Sparta P.D. Booking

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 00893

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/21/20 09:37

Subject Name

Test

Subject I.D.

RFI

Operator Name, I.D.

L. Nyström 615

Location

SPD Booking

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 00894

Temp Date Time ^{s/} 210L

Air Blank:
02/21/20 09:39 .000
Calibration Check:
24 02/21/20 09:39 .000

Subject Name

Test

Subject I.D.

Blank

Operator Name, I.D.

L. Nyström 615

Location

SPD Booking