



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY

**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN Nixa 107987	PRINTER SN Nixa 096.3580.875	DATE OF INSPECTION 02/04/2020
----------------------------------	---------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 715 W. Center Circle, Nixa	TIME OF INSPECTION 4:00 pm
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION                       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing Inc                      LOT # 19001                      EXP. DATE 03/12/2021

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0                      SIMULATOR SN SD 2731                      SIMULATOR EXP DATE 09/06/2020

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .100	TEST 2  .100	TEST 3  .100
--------------	--------------	--------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	1	(OVER .19)	3
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Meets or Exceeds Department of Health and Senior Services Rules and Regulations.

**INSPECTING OFFICER:**

SIGNATURE 	PRINT NAME Cpl. J. Barton
TYPE II PERMIT NUMBER/EXPIRATION DATE 290024      01/29/2021	TELEPHONE NUMBER (417) 725-2510

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

# Nixa Police Department

## Calibration Check slip's

AS IV Serial no: 107987  
Version no: 532B

TEST RECORD 01018

Temp Date Time <sup>9/</sup> 210L

Air Blank:  
02/04/20 16:03 .000  
Calibration:  
23 02/04/20 16:03 .100

Subject Name

TEST 1  
Subject I.D.

Operator Name, I.D.

[Signature] 403  
Location

AS IV Serial no: 107987  
Version no: 532B

TEST RECORD 01019

Temp Date Time <sup>9/</sup> 210L

Air Blank:  
02/04/20 16:05 .000  
Calibration Check:  
23 02/04/20 16:05 .100

Subject Name

TEST 2  
Subject I.D.

Operator Name, I.D.

[Signature] 403  
Location

AS IV Serial no: 107987  
Version no: 532B

TEST RECORD 01020

Temp Date Time <sup>9/</sup> 210L

Air Blank:  
02/04/20 16:07 .000  
Subject Test: Man  
25 02/04/20 16:07 .100

Subject Name

[Signature] 403  
Subject I.D.

Operator Name, I.D.

TEST 3  
Location

# Nixa Police Department

## RFI Evidence slip

AS IV Serial no: 107987  
Version no: 5028

TEST RECORD 01021

Temp: \_\_\_\_\_ Date: \_\_\_\_\_ Time: 210L  
g/

VOID: RFI  
12/02/04/20 10:00

Subject Name

**RFI**

Subject I.D.

Operator Name, I.D.

 403

Location

# Nixa Police Department

## Blank (Zero) test Evidence slip

AS IV Serial no: 187987  
Version no: 532R

TEST RECORD 01622

Temp Date Time 210L

Air Blank:  
R2/04/20 10:00 .000  
Subject Test: Auto  
25 R2/04/20 10:00 .000

Subject Name


BLANK

Subject I.D.

Operator Name I.D.

Locat No

403



RepCo Marketing Co  
3101-188 Stony Brook Drive  
Raleigh, NC 27604  
888-828-0227

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.**

**LOT NUMBER: 19001**

**EXPIRATION DATE: March 12, 2021** at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 19001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by accredited institution, Data Resources Inc., using NIST standards. Random samples were analyzed by Data Resources Inc. utilizing a gas chromatograph and found to contain .1210 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is March 13, 2019  
The expiration date for this lot number is March 12, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer, Operations Manager  
RepCo Marketing Co.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**JOSHUA C BARTON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/29/2019

NUMBER 290024

EXPIRES 1/29/2021

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BARTON, JOSHUA  
 Permit No 290024  
 Date Issued 1/29/2019 Date Expires 1/29/2021

