



**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107993	PRINTER SN 099.3586.836	DATE OF INSPECTION 10/27/2020
LOCATION OF INSTRUMENT (STREET AND CITY) 9623 SAINT CHARLES ROCK ROAD, BRECKENRIDGE HILLS POLICE DEPARTMENT		TIME OF INSPECTION 7:24 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) 20°C
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER GUTH LABORATORIES, INC. LOT # 20190 EXP. DATE 04/06/2022
- SIMULATOR TEMPERATURE ( $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$ ) 34.0 SIMULATOR SN SD2309 SIMULATOR EXP DATE 03/02/2021
- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within  $\pm 5\%$  of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <u>.101</u>	TEST 2 <u>.100</u>	TEST 3 <u>.099</u>
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	<u>0</u>	(.0-.04)	(.05-.09)	<u>1</u>	(.10-.14)	(.15-.19)	(OVER .19)	<u>2</u>
----------	----------	----------	-----------	----------	-----------	-----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE <i>Michael Presson</i>	PRINT NAME MICHAEL PRESSON
TYPE II PERMIT NUMBER/EXPIRATION DATE 290263/10-31-2021	TELEPHONE NUMBER (314) 426-1214

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS 10 Serial no: 107993  
Version no: 502B

TEST REPORT 00765

Temp Date Time 210L

Air Blank:  
10/27/20 19:24 .000  
Calibration Check:  
20 10/27/20 19:24 .101

Subject Name

Subject I.D.

Operator Name: I.D.

*Productive* *JS*  
Location

9655 Mo 180

GS 10 Serial no: 107993  
Version No: 532B

TEST RECORD 00766

Temp Date Time 210L

Air Blank:  
18/27/20 19:26 .000

Calibration Check:  
21 18/27/20 19:26 .100

Subject Name

Subject I.D.

Operator Name I.D.

*Podiatrist [Signature]* 734

Location

9628 mo 180

AS 10 Serial no: 107998  
Version no: 502B

TEST RECORD 00767

Temp Date Time 3/ 216L

Air Blank  
10/27/20 19:27 .000  
Calibration Check  
22 10/27/20 19:27 .000

Subject Name

Subject I.D.

Operator Name I.D.

*Polina...*

Location

*9623 MALSD*



AG ID Serial no: 147993  
Version no: 5328

LAST RECORD 00769

Name Date Time 210L

UNIT: FBI  
12-10-27-20 19:29

Subject Name

Subject I.D.

Operator Name: I.D.

*Richard J. ... 238*

Location

*9603 ...*



**GUTH LABORATORIES, INC.**

590 NORTH 87th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 20190 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on April 8, 2020, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is April 6, 2022 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN02271802 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

PERMIT  
TYPE II

MICHAEL B PRESSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 10/31/2019

NUMBER 290263

EXPIRES 10/31/2021

MO 580-0771 (8-10)

  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAG-4 (R6-10)

 STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator PRESSON, MICHAEL  
Permit No 290263  
Date Issued 10/31/2019 Date Expires 10/31/2021

