

RECEIVED

By Tracy Crews at 10:34 am, Mar 11, 2020


 MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107995	PRINTER SN 099.3586.789	DATE OF INSPECTION 03/10/2020
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Troop F Headquarters, 2920 Shamrock Drive, Jefferson City	TIME OF INSPECTION 9:24 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

 DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

 TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

 PRINTER WORKING PROPERLY

 TIME AND DATE DISPLAYING PROPERLY
BREATH ALCOHOL ACCURACY STANDARDS
 SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

 STANDARD SUPPLIER RepCo LOT # 19002 EXP. DATE 10/16/2021
 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIMULATOR SN MP2123 SIMULATOR EXP DATE 12/20/2020
 CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .103

TEST 2 .104

TEST 3 .105

 RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Calibrated time

INSPECTING OFFICER

SIGNATURE

PRINT NAME

M.A. Dickens

TYPE II PERMIT NUMBER/EXPIRATION DATE

290283 12/11/2021

TELEPHONE NUMBER

(573) 751-1000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 00646

Temp Date Time ^{s/} 210L

Air Blank:
03/10/20 21:24 .000
Calibration Check:
20 03/10/20 21:24 .103

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

MA Dickens 290283

Location

Troop F

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 00647

Temp Date Time ^{s/} 210L

Air Blank:
03/10/20 21:26 .000
Calibration Check:
21 03/10/20 21:26 .104

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

Dickens 290283

Location

Troop F

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 00648

Temp Date Time ^{s/} 210L

Air Blank:
03/10/20 21:30 .000
Calibration Check:
22 03/10/20 21:30 .105

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

Dickens 290283

Location

Troop F

AS IV Serial no: 107995
Version no: 532B

TEST RECORD 00649

Temp Date Time ^{s/} 210L

VOID: RFI
12 03/10/20 21:31

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

Dickens 290283

Location

Troop F

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 19002
EXPIRATION DATE: October 16, 2021 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 19002 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1231 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 17, 2019 The expiration date for this lot number is October 16, 2021 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

MARYLYN A DICKENS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/11/2018

NUMBER 280028

EXPIRES 1/11/2020

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator DICKENS, MARYLYN
 Permit No 280028
 Date Issued 1/11/2018 Date Expires 1/11/2020